

Continuing Calibration Report for SequenceID : FD022725AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049120.D

Aromatic C10-C12	6229425.000	40.000	4.076	5.794	155735.625	159020.606	2.066
Aromatic C12-C16	9514645.000	60.000	5.795	8.397	158577.417	160935.866	1.465
Aromatic C16-C21	11756074.000	80.000	8.398	12.661	146950.925	153359.309	4.179
Aromatic C21-C36	25098023.000	180.000	12.662	18.071	139433.461	143114.753	2.572
Aromatic EPH	52598167.000	360.000	4.076	18.071	146106.019	150128.823	2.680