

Continuing Calibration Report for SequenceID : FD022725AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049126.D

Aromatic C10-C12	6023865.000	40.000	4.076	5.794	150596.625	159020.606	5.297
Aromatic C12-C16	9168678.000	60.000	5.795	8.397	152811.300	160935.866	5.048
Aromatic C16-C21	11343539.000	80.000	8.398	12.661	141794.238	153359.309	7.541
Aromatic C21-C36	24057363.000	180.000	12.662	18.071	133652.017	143114.753	6.612
Aromatic EPH	50593445.000	360.000	4.076	18.071	140537.347	150128.823	6.389