

Continuing Calibration Report for SequenceID : FD022825AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049129.D

Aromatic C10-C12	6160252.000	40.000	4.076	5.794	154006.300	159020.606	3.153
Aromatic C12-C16	9402232.000	60.000	5.795	8.396	156703.867	160935.866	2.630
Aromatic C16-C21	11641914.000	80.000	8.397	12.661	145523.925	153359.309	5.109
Aromatic C21-C36	23657575.000	180.000	12.662	18.073	131430.972	143114.753	8.164
Aromatic EPH	50861973.000	360.000	4.076	18.073	141283.258	150128.823	5.892