

Continuing Calibration Report for SequenceID : FD022825AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049139.D

Aromatic C10-C12	6813477.000	40.000	4.076	5.794	170336.925	159020.606	-7.116
Aromatic C12-C16	10472510.000	60.000	5.795	8.396	174541.833	160935.866	-8.454
Aromatic C16-C21	13024891.000	80.000	8.397	12.661	162811.138	153359.309	-6.163
Aromatic C21-C36	26617558.000	180.000	12.662	18.073	147875.322	143114.753	-3.326
Aromatic EPH	56928436.000	360.000	4.076	18.073	158134.544	150128.823	-5.333