

Continuing Calibration Report for SequenceID : FD022825AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049149.D

Aromatic C10-C12	6973708.000	40.000	4.076	5.794	174342.700	159020.606	-9.635
Aromatic C12-C16	10519197.000	60.000	5.795	8.396	175319.950	160935.866	-8.938
Aromatic C16-C21	13182289.000	80.000	8.397	12.661	164778.613	153359.309	-7.446
Aromatic C21-C36	27715585.000	180.000	12.662	18.073	153975.472	143114.753	-7.589
Aromatic EPH	58390779.000	360.000	4.076	18.073	162196.608	150128.823	-8.038