

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1421 SAS No.: Q1421 SDG NO.: Q1421

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL01 Date Analyzed: 02/25/2025

Lab Sample No.: AR1660CCC500 Data File : PP070030.D Time Analyzed: 09:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.683	5.581	5.781	505.940	500.000	1.2
Aroclor-1016-2	5.704	5.602	5.802	520.740	500.000	4.1
Aroclor-1016-3	5.767	5.665	5.865	499.380	500.000	-0.1
Aroclor-1016-4	5.864	5.762	5.962	514.200	500.000	2.8
Aroclor-1016-5	6.157	6.056	6.256	514.170	500.000	2.8
Aroclor-1260-1	7.277	7.175	7.375	514.020	500.000	2.8
Aroclor-1260-2	7.530	7.429	7.629	496.310	500.000	-0.7
Aroclor-1260-3	7.884	7.788	7.988	585.370	500.000	17.1
Aroclor-1260-4	8.115	8.012	8.212	519.540	500.000	3.9
Aroclor-1260-5	8.435	8.333	8.533	515.640	500.000	3.1
Decachlorobiphenyl	10.262	10.156	10.356	51.550	50.000	3.1
Tetrachloro-m-xylene	4.529	4.427	4.627	53.390	50.000	6.8

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1421 SAS No.: Q1421 SDG NO.: Q1421

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL01 Date Analyzed: 02/25/2025

Lab Sample No.: AR1660CCC500 Data File : PP070030.D Time Analyzed: 09:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.920	4.820	5.020	511.860	500.000	2.4
Aroclor-1016-2	4.939	4.839	5.039	517.760	500.000	3.6
Aroclor-1016-3	5.117	5.017	5.217	535.770	500.000	7.2
Aroclor-1016-4	5.158	5.058	5.258	534.310	500.000	6.9
Aroclor-1016-5	5.374	5.273	5.473	527.970	500.000	5.6
Aroclor-1260-1	6.412	6.312	6.512	523.760	500.000	4.8
Aroclor-1260-2	6.599	6.500	6.700	520.310	500.000	4.1
Aroclor-1260-3	6.755	6.654	6.854	482.990	500.000	-3.4
Aroclor-1260-4	7.226	7.126	7.326	532.640	500.000	6.5
Aroclor-1260-5	7.467	7.367	7.567	503.450	500.000	0.7
Decachlorobiphenyl	8.889	8.788	8.988	48.530	50.000	-2.9
Tetrachloro-m-xylene	3.831	3.730	3.930	51.350	50.000	2.7