

### CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1421 SAS No.: Q1421 SDG NO.: Q1421

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL02 Date Analyzed: 02/25/2025

Lab Sample No.: AR1660CCC500 Data File : PP070045.D Time Analyzed: 15:27

| COMPOUND             | RT     | RT WINDOW<br>FROM TO |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|--------|----------------------|--------|--------------------|-------------------|------|
| Aroclor-1016-1       | 5.682  | 5.581                | 5.781  | 493.070            | 500.000           | -1.4 |
| Aroclor-1016-2       | 5.703  | 5.602                | 5.802  | 488.280            | 500.000           | -2.3 |
| Aroclor-1016-3       | 5.766  | 5.665                | 5.865  | 485.740            | 500.000           | -2.9 |
| Aroclor-1016-4       | 5.863  | 5.762                | 5.962  | 479.880            | 500.000           | -4.0 |
| Aroclor-1016-5       | 6.156  | 6.056                | 6.256  | 494.780            | 500.000           | -1.0 |
| Aroclor-1260-1       | 7.276  | 7.175                | 7.375  | 470.130            | 500.000           | -6.0 |
| Aroclor-1260-2       | 7.529  | 7.429                | 7.629  | 457.880            | 500.000           | -8.4 |
| Aroclor-1260-3       | 7.888  | 7.788                | 7.988  | 484.630            | 500.000           | -3.1 |
| Aroclor-1260-4       | 8.113  | 8.012                | 8.212  | 477.180            | 500.000           | -4.6 |
| Aroclor-1260-5       | 8.433  | 8.333                | 8.533  | 494.540            | 500.000           | -1.1 |
| Decachlorobiphenyl   | 10.258 | 10.156               | 10.356 | 49.540             | 50.000            | -0.9 |
| Tetrachloro-m-xylene | 4.528  | 4.427                | 4.627  | 50.690             | 50.000            | 1.4  |

### CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1421 SAS No.: Q1421 SDG NO.: Q1421

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL02 Date Analyzed: 02/25/2025

Lab Sample No.: AR1660CCC500 Data File : PP070045.D Time Analyzed: 15:27

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 4.922 | 4.820     | 5.020 | 474.050            | 500.000           | -5.2  |
| Aroclor-1016-2       | 4.941 | 4.839     | 5.039 | 475.050            | 500.000           | -5.0  |
| Aroclor-1016-3       | 5.118 | 5.017     | 5.217 | 482.530            | 500.000           | -3.5  |
| Aroclor-1016-4       | 5.160 | 5.058     | 5.258 | 484.710            | 500.000           | -3.1  |
| Aroclor-1016-5       | 5.375 | 5.273     | 5.473 | 481.080            | 500.000           | -3.8  |
| Aroclor-1260-1       | 6.413 | 6.312     | 6.512 | 468.300            | 500.000           | -6.3  |
| Aroclor-1260-2       | 6.601 | 6.500     | 6.700 | 476.360            | 500.000           | -4.7  |
| Aroclor-1260-3       | 6.755 | 6.654     | 6.854 | 428.530            | 500.000           | -14.3 |
| Aroclor-1260-4       | 7.228 | 7.126     | 7.326 | 444.280            | 500.000           | -11.1 |
| Aroclor-1260-5       | 7.468 | 7.367     | 7.567 | 467.420            | 500.000           | -6.5  |
| Decachlorobiphenyl   | 8.890 | 8.788     | 8.988 | 46.920             | 50.000            | -6.2  |
| Tetrachloro-m-xylene | 3.832 | 3.730     | 3.930 | 48.180             | 50.000            | -3.6  |