

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022425\
 Data File : PP070020.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Feb 2025 21:29
 Operator : YP\AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC750

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/25/2025
 Supervised By :Ankita Jodhani 02/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 25 04:26:41 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 25 04:25:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.526	3.830	103.2E6	64166467	67.333	64.432m
2) SA Decachlor...	10.254	8.887	137.1E6	125.3E6	71.547	74.298
Target Compounds						
41) L9 AR-1268-1	8.746	7.750	138.6E6	118.3E6	715.609	741.494
42) L9 AR-1268-2	8.840	7.815	119.7E6	102.9E6	711.321	749.564
43) L9 AR-1268-3	9.073	8.022	102.9E6	86951030	707.478	721.763
44) L9 AR-1268-4	9.492	8.317	45213739	37616870	702.540	710.904
45) L9 AR-1268-5	9.911	8.622	301.0E6	264.8E6	719.439	769.421

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022425\
Data File : PP070020.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Feb 2025 21:29
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1268ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/25/2025
Supervised By :Ankita Jodhani 02/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 25 04:26:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 25 04:25:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.526 min
Delta R.T.: 0.000 min
Response: 103201349
Conc: 67.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/25/2025
Supervised By :Ankita Jodhani 02/25/2025




#1 Tetrachloro-m-xylene

R.T.: 3.830 min
Delta R.T.: 0.000 min
Response: 64166467
Conc: 64.43 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.254 min
Delta R.T.: 0.000 min
Response: 137120935
Conc: 71.55 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.887 min
Delta R.T.: 0.000 min
Response: 125324212
Conc: 74.30 ng/ml




#41 AR-1268-1

R.T.: 8.746 min
Delta R.T.: 0.000 min
Response: 138562138
Conc: 715.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/25/2025
Supervised By :Ankita Jodhani 02/25/2025




#41 AR-1268-1

R.T.: 7.750 min
Delta R.T.: 0.000 min
Response: 118319583
Conc: 741.49 ng/ml




#42 AR-1268-2

R.T.: 8.840 min
Delta R.T.: 0.000 min
Response: 119720560
Conc: 711.32 ng/ml



#42 AR-1268-2

R.T.: 7.815 min
Delta R.T.: 0.000 min
Response: 102918371
Conc: 749.56 ng/ml




#43 AR-1268-3

R.T.: 9.073 min
Delta R.T.: 0.000 min
Response: 102857887
Conc: 707.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/25/2025
Supervised By :Ankita Jodhani 02/25/2025




#43 AR-1268-3

R.T.: 8.022 min
Delta R.T.: 0.000 min
Response: 86951030
Conc: 721.76 ng/ml




#44 AR-1268-4

R.T.: 9.492 min
Delta R.T.: 0.000 min
Response: 45213739
Conc: 702.54 ng/ml



#44 AR-1268-4

R.T.: 8.317 min
Delta R.T.: 0.001 min
Response: 37616870
Conc: 710.90 ng/ml




#45 AR-1268-5

R.T.: 9.911 min
Delta R.T.: 0.000 min
Response: 300965315
Conc: 719.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/25/2025
Supervised By :Ankita Jodhani 02/25/2025



#45 AR-1268-5

R.T.: 8.622 min
Delta R.T.: 0.000 min
Response: 264796784
Conc: 769.42 ng/ml