



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1421
Client: Weston Solutions, Inc.

Order ID: Q1421
Project ID: RFP 905

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :								

Total Concentration: 0.000



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	02/24/25	
Project:	RFP 905		Date Received:	02/24/25	
Client Sample ID:	P001-CLAY-CF01-01		SDG No.:	Q1421	
Lab Sample ID:	Q1421-01		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	70.3	Decanted:
Sample Wt/Vol:	30.04	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP070047.D	1	02/25/25 09:02	02/25/25 16:00	PB166854

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.80	U	4.80	24.1	ug/kg
11104-28-2	Aroclor-1221	9.10	U	9.10	24.1	ug/kg
11141-16-5	Aroclor-1232	4.80	U	4.80	24.1	ug/kg
53469-21-9	Aroclor-1242	4.80	U	4.80	24.1	ug/kg
12672-29-6	Aroclor-1248	11.2	U	11.2	24.1	ug/kg
11097-69-1	Aroclor-1254	3.90	U	3.90	24.1	ug/kg
37324-23-5	Aroclor-1262	6.50	U	6.50	24.1	ug/kg
11100-14-4	Aroclor-1268	4.90	U	4.90	24.1	ug/kg
11096-82-5	Aroclor-1260	4.10	U	4.10	24.1	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.4		32 - 144	107%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.8		32 - 175	99%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	02/24/25	
Project:	RFP 905		Date Received:	02/24/25	
Client Sample ID:	P001-CLAY-CF01-02		SDG No.:	Q1421	
Lab Sample ID:	Q1421-07		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	75.2	Decanted:
Sample Wt/Vol:	30.07	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP070050.D	1	02/25/25 09:02	02/25/25 16:48	PB166854

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.50	U	4.50	22.6	ug/kg
11104-28-2	Aroclor-1221	8.50	U	8.50	22.6	ug/kg
11141-16-5	Aroclor-1232	4.50	U	4.50	22.6	ug/kg
53469-21-9	Aroclor-1242	4.50	U	4.50	22.6	ug/kg
12672-29-6	Aroclor-1248	10.5	U	10.5	22.6	ug/kg
11097-69-1	Aroclor-1254	3.60	U	3.60	22.6	ug/kg
37324-23-5	Aroclor-1262	6.10	U	6.10	22.6	ug/kg
11100-14-4	Aroclor-1268	4.60	U	4.60	22.6	ug/kg
11096-82-5	Aroclor-1260	3.90	U	3.90	22.6	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	27.5		32 - 144	137%	SPK: 20
2051-24-3	Decachlorobiphenyl	22.7		32 - 175	113%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	02/24/25	
Project:	RFP 905		Date Received:	02/24/25	
Client Sample ID:	P001-CLAY-CF02-01		SDG No.:	Q1421	
Lab Sample ID:	Q1421-09		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	78.4	Decanted:
Sample Wt/Vol:	30.02	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP070051.D	1	02/25/25 09:02	02/25/25 17:05	PB166854

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.30	U	4.30	21.7	ug/kg
11104-28-2	Aroclor-1221	8.20	U	8.20	21.7	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	21.7	ug/kg
53469-21-9	Aroclor-1242	4.30	U	4.30	21.7	ug/kg
12672-29-6	Aroclor-1248	10.1	U	10.1	21.7	ug/kg
11097-69-1	Aroclor-1254	3.50	U	3.50	21.7	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	21.7	ug/kg
11100-14-4	Aroclor-1268	4.40	U	4.40	21.7	ug/kg
11096-82-5	Aroclor-1260	3.70	U	3.70	21.7	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	27.9		32 - 144	140%	SPK: 20
2051-24-3	Decachlorobiphenyl	24.4		32 - 175	122%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

LAB CHRONICLE

OrderID:	Q1421	OrderDate:	2/24/2025 2:56:42 PM
Client:	Weston Solutions, Inc.	Project:	RFP 905
Contact:	Smita Sumbaly	Location:	H31,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1421-01	P001-CLAY-CF01-01	SOIL	PCB	8082A	02/24/25	02/25/25	02/25/25	02/24/25
Q1421-07	P001-CLAY-CF01-02	SOIL	PCB	8082A	02/24/25	02/25/25	02/25/25	02/24/25
Q1421-09	P001-CLAY-CF02-01	SOIL	PCB	8082A	02/24/25	02/25/25	02/25/25	02/24/25