

Surrogate Summary

SDG No.: Q1421

Client: Weston Solutions, Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP069995.D	PIBLK-PP069995.D	Tetrachloro-m-xylene	1	20	21.9	109		60	140
		Decachlorobiphenyl	1	20	21.8	109		60	140
		Tetrachloro-m-xylene	2	20	22.1	110		60	140
		Decachlorobiphenyl	2	20	21.6	108		60	140
I.BLK-PP070034.D	PIBLK-PP070034.D	Tetrachloro-m-xylene	1	20	16.9	85		60	140
		Decachlorobiphenyl	1	20	17.4	87		60	140
		Tetrachloro-m-xylene	2	20	16.8	84		60	140
		Decachlorobiphenyl	2	20	18.0	90		60	140
PB166854BL	PB166854BL	Tetrachloro-m-xylene	1	20	23.9	119		32	144
		Decachlorobiphenyl	1	20	24.3	122		32	175
		Tetrachloro-m-xylene	2	20	24.1	120		32	144
		Decachlorobiphenyl	2	20	23.1	116		32	175
PB166854BS	PB166854BS	Tetrachloro-m-xylene	1	20	26.0	130		32	144
		Decachlorobiphenyl	1	20	25.4	127		32	175
		Tetrachloro-m-xylene	2	20	24.9	125		32	144
		Decachlorobiphenyl	2	20	24.7	123		32	175
I.BLK-PP070046.D	PIBLK-PP070046.D	Tetrachloro-m-xylene	1	20	17.7	89		60	140
		Decachlorobiphenyl	1	20	17.4	87		60	140
		Tetrachloro-m-xylene	2	20	18.1	91		60	140
		Decachlorobiphenyl	2	20	18.0	90		60	140
Q1421-01	P001-CLAY-CF01-01	Tetrachloro-m-xylene	1	20	21.1	105		32	144
		Decachlorobiphenyl	1	20	19.1	96		32	175
		Tetrachloro-m-xylene	2	20	21.4	107		32	144
		Decachlorobiphenyl	2	20	19.8	99		32	175
Q1421-02MS	P001-CLAY-CF01-01MS	Tetrachloro-m-xylene	1	20	28.0	140		32	144
		Decachlorobiphenyl	1	20	25.6	128		32	175
		Tetrachloro-m-xylene	2	20	27.5	137		32	144
		Decachlorobiphenyl	2	20	25.2	126		32	175
Q1421-03MSD	P001-CLAY-CF01-01MSD	Tetrachloro-m-xylene	1	20	28.0	140		32	144
		Decachlorobiphenyl	1	20	26.0	130		32	175
		Tetrachloro-m-xylene	2	20	25.9	130		32	144
		Decachlorobiphenyl	2	20	25.2	126		32	175
Q1421-07	P001-CLAY-CF01-02	Tetrachloro-m-xylene	1	20	26.4	132		32	144
		Decachlorobiphenyl	1	20	22.7	113		32	175
		Tetrachloro-m-xylene	2	20	27.5	137		32	144
		Decachlorobiphenyl	2	20	22.0	110		32	175
Q1421-09	P001-CLAY-CF02-01	Tetrachloro-m-xylene	1	20	26.2	131		32	144
		Decachlorobiphenyl	1	20	23.5	118		32	175
		Tetrachloro-m-xylene	2	20	27.9	140		32	144
		Decachlorobiphenyl	2	20	24.4	122		32	175
I.BLK-PP070058.D	PIBLK-PP070058.D	Tetrachloro-m-xylene	1	20	17.8	89		60	140

Surrogate Summary

SDG No.: Q1421

Client: Weston Solutions, Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP070058.D	PIBLK-PP070058.D	Decachlorobiphenyl	1	20	17.8	89		60	140
		Tetrachloro-m-xylene	2	20	18.6	93		60	140
		Decachlorobiphenyl	2	20	19.8	99		60	140