

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1421

Client: Weston Solutions, Inc.

Analytical Method: 8082A

DataFile : PP070049.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	P001-CLAY-CF01-01MSD											
Q1421-03MSD	AR1016	236.8	0	249	ug/kg	105		1		55	146	20
	AR1260	236.8	0	226	ug/kg	95		0		45	144	20