

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1421

Client: Weston Solutions, Inc.

Analytical Method: 8081B

DataFile : PL094393.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	P001-CLAY-CF01-01MS											
Q1421-02MS	alpha-BHC	23.66	0	23.9	ug/kg	101				60	144	
	beta-BHC	23.66	0	24.5	ug/kg	104				54	143	
	delta-BHC	23.66	0	23.9	ug/kg	101				29	151	
	gamma-BHC (Lindane)	23.66	0	23.3	ug/kg	98				61	140	
	Heptachlor	23.66	0	25.0	ug/kg	106				63	135	
	Aldrin	23.66	0	23.0	ug/kg	97				49	139	
	Heptachlor epoxide	23.66	0	23.6	ug/kg	100				41	156	
	Endosulfan I	23.66	0	24.0	ug/kg	101				56	142	
	Dieldrin	23.66	0	23.6	ug/kg	100				47	161	
	4,4'-DDE	23.66	0	24.1	ug/kg	102				55	136	
	Endrin	23.66	0	26.4	ug/kg	112				57	139	
	Endosulfan II	23.66	0	24.5	ug/kg	104				40	163	
	4,4'-DDD	23.66	0	26.7	ug/kg	113				47	163	
	Endosulfan sulfate	23.66	0	23.2	ug/kg	98				62	139	
	4,4'-DDT	23.66	0	23.2	ug/kg	98				51	146	
	Methoxychlor	23.66	0	22.6	ug/kg	96				54	136	
	Endrin ketone	23.66	0	22.8	ug/kg	96				60	129	
	Endrin aldehyde	23.66	0	22.0	ug/kg	93				59	132	
	alpha-Chlordane	23.66	0	23.9	ug/kg	101				39	166	
	gamma-Chlordane	23.66	0	24.1	ug/kg	102				44	175	