

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1421

Client: Weston Solutions, Inc.

Analytical Method: 8081B

DataFile : PL094394.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	P001-CLAY-CF01-01MSD											
Q1421-03MSD	alpha-BHC	23.68	0	24.0	ug/kg	101		0		60	144	20
	beta-BHC	23.68	0	25.5	ug/kg	108		4		54	143	20
	delta-BHC	23.68	0	23.6	ug/kg	100		1		29	151	20
	gamma-BHC (Lindane)	23.68	0	23.6	ug/kg	100		2		61	140	20
	Heptachlor	23.68	0	24.0	ug/kg	101		5		63	135	20
	Aldrin	23.68	0	23.4	ug/kg	99		2		49	139	20
	Heptachlor epoxide	23.68	0	23.8	ug/kg	101		1		41	156	20
	Endosulfan I	23.68	0	24.3	ug/kg	103		2		56	142	20
	Dieldrin	23.68	0	24.1	ug/kg	102		2		47	161	20
	4,4'-DDE	23.68	0	24.7	ug/kg	104		2		55	136	20
	Endrin	23.68	0	27.2	ug/kg	115		3		57	139	20
	Endosulfan II	23.68	0	24.6	ug/kg	104		0		40	163	20
	4,4'-DDD	23.68	0	27.2	ug/kg	115		2		47	163	20
	Endosulfan sulfate	23.68	0	23.6	ug/kg	100		2		62	139	20
	4,4'-DDT	23.68	0	24.0	ug/kg	101		3		51	146	20
	Methoxychlor	23.68	0	22.6	ug/kg	95		1		54	136	20
	Endrin ketone	23.68	0	23.8	ug/kg	101		5		60	129	20
	Endrin aldehyde	23.68	0	23.2	ug/kg	98		5		59	132	20
	alpha-Chlordane	23.68	0	24.3	ug/kg	103		2		39	166	20
	gamma-Chlordane	23.68	0	24.5	ug/kg	103		1		44	175	20