

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1421 SAS No.: Q1421 SDG NO.: Q1421

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL01 Date Analyzed: 02/27/2025

Lab Sample No.: AR1660CCC500 Data File : PP070094.D Time Analyzed: 16:39

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.681	5.581	5.781	447.550	500.000	-10.5
Aroclor-1016-2	5.703	5.602	5.802	455.770	500.000	-8.8
Aroclor-1016-3	5.765	5.665	5.865	447.400	500.000	-10.5
Aroclor-1016-4	5.863	5.762	5.962	473.180	500.000	-5.4
Aroclor-1016-5	6.156	6.056	6.256	457.790	500.000	-8.4
Aroclor-1260-1	7.275	7.175	7.375	458.170	500.000	-8.4
Aroclor-1260-2	7.529	7.429	7.629	427.130	500.000	-14.6
Aroclor-1260-3	7.888	7.788	7.988	465.570	500.000	-6.9
Aroclor-1260-4	8.112	8.012	8.212	434.870	500.000	-13.0
Aroclor-1260-5	8.434	8.333	8.533	461.330	500.000	-7.7
Decachlorobiphenyl	10.259	10.156	10.356	46.890	50.000	-6.2
Tetrachloro-m-xylene	4.528	4.427	4.627	52.870	50.000	5.7

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1421 SAS No.: Q1421 SDG NO.: Q1421

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL01 Date Analyzed: 02/27/2025

Lab Sample No.: AR1660CCC500 Data File : PP070094.D Time Analyzed: 16:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.922	4.820	5.020	525.870	500.000	5.2
Aroclor-1016-2	4.941	4.839	5.039	523.420	500.000	4.7
Aroclor-1016-3	5.118	5.017	5.217	541.660	500.000	8.3
Aroclor-1016-4	5.159	5.058	5.258	533.560	500.000	6.7
Aroclor-1016-5	5.375	5.273	5.473	537.410	500.000	7.5
Aroclor-1260-1	6.413	6.312	6.512	484.960	500.000	-3.0
Aroclor-1260-2	6.601	6.500	6.700	492.120	500.000	-1.6
Aroclor-1260-3	6.756	6.654	6.854	462.810	500.000	-7.4
Aroclor-1260-4	7.228	7.126	7.326	471.760	500.000	-5.6
Aroclor-1260-5	7.469	7.367	7.567	483.810	500.000	-3.2
Decachlorobiphenyl	8.890	8.788	8.988	44.550	50.000	-10.9
Tetrachloro-m-xylene	3.833	3.730	3.930	51.970	50.000	3.9