

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1421 SAS No.: Q1421 SDG NO.: Q1421

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL02 Date Analyzed: 02/27/2025

Lab Sample No.: AR1660CCC500 Data File : PP070109.D Time Analyzed: 21:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.682	5.581	5.781	497.000	500.000	-0.6
Aroclor-1016-2	5.704	5.602	5.802	522.380	500.000	4.5
Aroclor-1016-3	5.767	5.665	5.865	503.190	500.000	0.6
Aroclor-1016-4	5.864	5.762	5.962	517.500	500.000	3.5
Aroclor-1016-5	6.157	6.056	6.256	520.840	500.000	4.2
Aroclor-1260-1	7.277	7.175	7.375	542.620	500.000	8.5
Aroclor-1260-2	7.531	7.429	7.629	501.390	500.000	0.3
Aroclor-1260-3	7.889	7.788	7.988	528.550	500.000	5.7
Aroclor-1260-4	8.113	8.012	8.212	510.190	500.000	2.0
Aroclor-1260-5	8.435	8.333	8.533	511.770	500.000	2.4
Decachlorobiphenyl	10.259	10.156	10.356	50.950	50.000	1.9
Tetrachloro-m-xylene	4.529	4.427	4.627	53.660	50.000	7.3

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1421 SAS No.: Q1421 SDG NO.: Q1421

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL02 Date Analyzed: 02/27/2025

Lab Sample No.: AR1660CCC500 Data File : PP070109.D Time Analyzed: 21:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.921	4.820	5.020	519.430	500.000	3.9
Aroclor-1016-2	4.939	4.839	5.039	514.720	500.000	2.9
Aroclor-1016-3	5.117	5.017	5.217	542.730	500.000	8.5
Aroclor-1016-4	5.159	5.058	5.258	533.660	500.000	6.7
Aroclor-1016-5	5.374	5.273	5.473	559.180	500.000	11.8
Aroclor-1260-1	6.412	6.312	6.512	517.680	500.000	3.5
Aroclor-1260-2	6.600	6.500	6.700	518.020	500.000	3.6
Aroclor-1260-3	6.754	6.654	6.854	484.090	500.000	-3.2
Aroclor-1260-4	7.226	7.126	7.326	519.730	500.000	3.9
Aroclor-1260-5	7.468	7.367	7.567	518.470	500.000	3.7
Decachlorobiphenyl	8.889	8.788	8.988	49.930	50.000	-0.1
Tetrachloro-m-xylene	3.832	3.730	3.930	54.710	50.000	9.4