

Surrogate Summary

SDG No.: Q1421

Client: Weston Solutions, Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP069995.D	PIBLK-PP069995.D	Tetrachloro-m-xylene	1	20	21.9	109		60	140
		Decachlorobiphenyl	1	20	21.8	109		60	140
		Tetrachloro-m-xylene	2	20	22.1	110		60	140
		Decachlorobiphenyl	2	20	21.6	108		60	140
I.BLK-PP070098.D	PIBLK-PP070098.D	Tetrachloro-m-xylene	1	20	18.8	94		60	140
		Decachlorobiphenyl	1	20	17.6	88		60	140
		Tetrachloro-m-xylene	2	20	19.7	98		60	140
		Decachlorobiphenyl	2	20	18.4	92		60	140
PB166895BL	PB166895BL	Tetrachloro-m-xylene	1	20	22.5	113		10	157
		Decachlorobiphenyl	1	20	21.3	107		10	173
		Tetrachloro-m-xylene	2	20	23.6	118		10	157
		Decachlorobiphenyl	2	20	21.9	110		10	173
PB166895BS	PB166895BS	Tetrachloro-m-xylene	1	20	23.7	119		10	157
		Decachlorobiphenyl	1	20	21.8	109		10	173
		Tetrachloro-m-xylene	2	20	23.7	119		10	157
		Decachlorobiphenyl	2	20	22.7	114		10	173
PB166895TB	PB166895TB	Tetrachloro-m-xylene	1	20	22.8	114		10	157
		Decachlorobiphenyl	1	20	21.8	109		10	173
		Tetrachloro-m-xylene	2	20	23.8	119		10	157
		Decachlorobiphenyl	2	20	22.1	110		10	173
Q1421-04	P001-CLAY-CF01-01	Tetrachloro-m-xylene	1	20	25.6	128		10	157
		Decachlorobiphenyl	1	20	24.7	123		10	173
		Tetrachloro-m-xylene	2	20	26.3	131		10	157
		Decachlorobiphenyl	2	20	23.8	119		10	173
Q1421-05MS	P001-CLAY-CF01-01MS	Tetrachloro-m-xylene	1	20	26.1	131		10	157
		Decachlorobiphenyl	1	20	23.2	116		10	173
		Tetrachloro-m-xylene	2	20	24.9	124		10	157
		Decachlorobiphenyl	2	20	23.5	118		10	173
Q1421-06MSD	P001-CLAY-CF01-01MSD	Tetrachloro-m-xylene	1	20	26.1	131		10	157
		Decachlorobiphenyl	1	20	23.3	116		10	173
		Tetrachloro-m-xylene	2	20	25.1	125		10	157
		Decachlorobiphenyl	2	20	22.8	114		10	173
Q1421-08	P001-CLAY-CF01-02	Tetrachloro-m-xylene	1	20	24.8	124		10	157
		Decachlorobiphenyl	1	20	22.9	115		10	173
		Tetrachloro-m-xylene	2	20	24.8	124		10	157
		Decachlorobiphenyl	2	20	21.9	110		10	173
Q1421-10	P001-CLAY-CF02-01	Tetrachloro-m-xylene	1	20	23.4	117		10	157
		Decachlorobiphenyl	1	20	20.9	104		10	173
		Tetrachloro-m-xylene	2	20	22.9	114		10	157
		Decachlorobiphenyl	2	20	20.9	104		10	173
I.BLK-PP070113.D	PIBLK-PP070113.D	Tetrachloro-m-xylene	1	20	19.3	96		60	140

Surrogate Summary

SDG No.: Q1421

Client: Weston Solutions, Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP070113.D	PIBLK-PP070113.D	Decachlorobiphenyl	1	20	18.6	93		60	140
		Tetrachloro-m-xylene	2	20	19.9	99		60	140
		Decachlorobiphenyl	2	20	19.2	96		60	140