

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: ROYF02

Lab Code: CHEM **Case No.:** Q1421 **SAS No.:** Q1421 **SDG NO.:** Q1421

Instrument ID: ECD_L **Calibration Date(s):** 02/21/2025 02/21/2025
Calibration Times: 10:56 11:51

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 100 = <u>PL094326.D</u>		CF 075 = <u>PL094327.D</u>			
CF 050 = <u>PL094328.D</u>		CF 025 = <u>PL094329.D</u>		CF 005 = <u>PL094330.D</u>			
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	1803580000	1795660000	1810310000	1935150000	1904070000	1849750000	4
4,4'-DDE	2493500000	2470740000	2506610000	2704950000	2882800000	2611720000	7
4,4'-DDT	1974760000	1979910000	2019710000	2146120000	2379640000	2100030000	8
Aldrin	3281770000	3217010000	3255180000	3477160000	3927610000	3431750000	9
alpha-BHC	3748810000	3649740000	3623010000	3780620000	4111130000	3782660000	5
alpha-Chlordane	2727770000	2693090000	2738980000	2936760000	3388830000	2897090000	10
beta-BHC	1501030000	1494420000	1485840000	1667440000	1835770000	1596900000	10
Decachlorobiphenyl	1933170000	1963230000	2030640000	2203310000	2505970000	2127260000	11
delta-BHC	3482300000	3394730000	3353250000	3521830000	3939260000	3538270000	7
Dieldrin	2692690000	2671770000	2693570000	2901890000	3285250000	2849030000	9
Endosulfan I	2547690000	2536030000	2605780000	2783760000	3269880000	2748630000	11
Endosulfan II	2275250000	2276760000	2334150000	2535820000	3164240000	2517240000	15
Endosulfan sulfate	2066530000	2082580000	2124320000	2325240000	2720550000	2263840000	12
Endrin	2293020000	2294250000	2335150000	2563130000	3275860000	2552280000	16
Endrin aldehyde	1783210000	1799570000	1847420000	2020050000	2283730000	1946800000	11
Endrin ketone	2341850000	2345380000	2389970000	2583410000	2873950000	2506910000	9
gamma-BHC (Lindane)	3636600000	3544540000	3535300000	3710470000	4046760000	3694740000	6
gamma-Chlordane	2749680000	2723950000	2753190000	2951510000	3502350000	2936140000	11
Heptachlor	3238030000	3195300000	3207600000	3444620000	3866240000	3390360000	8
Heptachlor epoxide	2854740000	2830940000	2868110000	3075040000	3438100000	3013390000	9
Methoxychlor	1034140000	1042330000	1072820000	1161190000	1244730000	1111040000	8
Tetrachloro-m-xylene	2440810000	2429970000	2443520000	2602410000	2778780000	2539100000	6

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: ROYF02

Lab Code: CHEM **Case No.:** Q1421 **SAS No.:** Q1421 **SDG NO.:** Q1421

Instrument ID: ECD_L **Calibration Date(s):** 02/21/2025 02/21/2025
Calibration Times: 10:56 11:51

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 100 = <u>PL094326.D</u>		CF 075 = <u>PL094327.D</u>			
CF 050 = <u>PL094328.D</u>		CF 025 = <u>PL094329.D</u>		CF 005 = <u>PL094330.D</u>			
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	3335160000	3215540000	3139450000	3166740000	3069690000	3185310000	3
4,4'-DDE	4281660000	4182440000	4118800000	4213820000	4203280000	4200000000	1
4,4'-DDT	3703100000	3584290000	3529700000	3537620000	3430330000	3557010000	3
Aldrin	4820710000	4668200000	4536250000	4555960000	4520730000	4620370000	3
alpha-BHC	5213240000	5010060000	4886090000	4788800000	4675760000	4914790000	4
alpha-Chlordane	4373480000	4258810000	4191540000	4303990000	4460210000	4317600000	2
beta-BHC	1965110000	1930080000	1919640000	1976280000	2153520000	1988920000	5
Decachlorobiphenyl	3675900000	3638830000	3681980000	3820210000	4411300000	3845640000	8
delta-BHC	5032580000	4845250000	4682410000	4626330000	4491990000	4735710000	4
Dieldrin	4518600000	4388970000	4327740000	4347710000	4325860000	4381780000	2
Endosulfan I	4036260000	3934540000	3904680000	3979590000	4077800000	3986580000	2
Endosulfan II	3796570000	3715310000	3681290000	3767040000	3737420000	3739530000	1
Endosulfan sulfate	3706520000	3626300000	3609280000	3706820000	3819330000	3693650000	2
Endrin	3538690000	3451940000	3434060000	3498850000	3100040000	3404720000	5
Endrin aldehyde	3123920000	3068890000	3064770000	3158970000	3385600000	3160430000	4
Endrin ketone	4300600000	4199470000	4213040000	4281500000	4343010000	4267520000	1
gamma-BHC (Lindane)	4983380000	4798550000	4658170000	4606370000	4503230000	4709940000	4
gamma-Chlordane	4438420000	4307340000	4231380000	4301670000	4415360000	4338830000	2
Heptachlor	4844080000	4715390000	4629500000	4679640000	4732940000	4720310000	2
Heptachlor epoxide	4316300000	4215090000	4165700000	4257130000	4322680000	4255380000	2
Methoxychlor	1924140000	1903960000	1935750000	2006380000	1984890000	1951030000	2
Tetrachloro-m-xylene	3209080000	3135420000	3127500000	3185970000	3297890000	3191170000	2