

Surrogate Summary

SDG No.: Q1421

Client: Weston Solutions, Inc.

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL094323.D	PIBLK-PL094323.D	Decachlorobiphenyl	1	20	24.7	123		43	140
		Tetrachloro-m-xylene	1	20	23.3	116		77	126
		Decachlorobiphenyl	2	20	23.6	118		43	140
		Tetrachloro-m-xylene	2	20	22.8	114		77	126
I.BLK-PL094432.D	PIBLK-PL094432.D	Decachlorobiphenyl	1	20	19.8	99		43	140
		Tetrachloro-m-xylene	1	20	20.9	105		77	126
		Decachlorobiphenyl	2	20	18.9	94		43	140
		Tetrachloro-m-xylene	2	20	20.8	104		77	126
PB166896BL	PB166896BL	Decachlorobiphenyl	1	20	24.2	121		43	140
		Tetrachloro-m-xylene	1	20	24.2	121		77	126
		Decachlorobiphenyl	2	20	24.2	121		43	140
		Tetrachloro-m-xylene	2	20	23.2	116		77	126
PB166896BS	PB166896BS	Decachlorobiphenyl	1	20	20.2	101		43	140
		Tetrachloro-m-xylene	1	20	19.9	99		77	126
		Decachlorobiphenyl	2	20	19.7	98		43	140
		Tetrachloro-m-xylene	2	20	19.3	97		77	126
PB166896TB	PB166896TB	Decachlorobiphenyl	1	20	24.2	121		43	140
		Tetrachloro-m-xylene	1	20	24.2	121		77	126
		Decachlorobiphenyl	2	20	24.0	120		43	140
		Tetrachloro-m-xylene	2	20	23.9	120		77	126
Q1421-04	P001-CLAY-CF01-01	Decachlorobiphenyl	1	20	25.2	126		43	140
		Tetrachloro-m-xylene	1	20	25.1	126		77	126
		Decachlorobiphenyl	2	20	25.0	125		43	140
		Tetrachloro-m-xylene	2	20	25.4	127	*	77	126
Q1421-05MS	P001-CLAY-CF01-01MS	Decachlorobiphenyl	1	20	25.2	126		43	140
		Tetrachloro-m-xylene	1	20	26.3	132	*	77	126
		Decachlorobiphenyl	2	20	25.4	127		43	140
		Tetrachloro-m-xylene	2	20	25.4	127	*	77	126
Q1421-06MSD	P001-CLAY-CF01-01MSD	Decachlorobiphenyl	1	20	25.7	128		43	140
		Tetrachloro-m-xylene	1	20	25.7	129	*	77	126
		Decachlorobiphenyl	2	20	25.8	129		43	140
		Tetrachloro-m-xylene	2	20	25.8	129	*	77	126
Q1421-08	P001-CLAY-CF01-02	Decachlorobiphenyl	1	20	25.6	128		43	140
		Tetrachloro-m-xylene	1	20	27.1	136	*	77	126
		Decachlorobiphenyl	2	20	25.4	127		43	140
		Tetrachloro-m-xylene	2	20	27.1	135	*	77	126
Q1421-10	P001-CLAY-CF02-01	Decachlorobiphenyl	1	20	25.4	127		43	140
		Tetrachloro-m-xylene	1	20	21.1	106		77	126
		Decachlorobiphenyl	2	20	24.3	122		43	140
		Tetrachloro-m-xylene	2	20	23.9	120		77	126
I.BLK-PL094443.D	PIBLK-PL094443.D	Decachlorobiphenyl	1	20	20.2	101		43	140

Surrogate Summary

SDG No.: Q1421

Client: Weston Solutions, Inc.

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL094443.D	PIBLK-PL094443.D	Tetrachloro-m-xylene	1	20	20.8	104		77	126
		Decachlorobiphenyl	2	20	18.8	94		43	140
I.BLK-PL094459.D	PIBLK-PL094459.D	Tetrachloro-m-xylene	2	20	20.4	102		77	126
		Decachlorobiphenyl	1	20	18.7	93		43	140
		Tetrachloro-m-xylene	1	20	20.1	100		77	126
		Decachlorobiphenyl	2	20	15.8	79		43	140
		Tetrachloro-m-xylene	2	20	20.6	103		77	126
Q1421-08RE	P001-CLAY-CF01-02RE	Decachlorobiphenyl	1	20	23.8	119		43	140
		Tetrachloro-m-xylene	1	20	26.4	132	*	77	126
		Decachlorobiphenyl	2	20	22.3	111		43	140
		Tetrachloro-m-xylene	2	20	26.7	134	*	77	126
I.BLK-PL094462.D	PIBLK-PL094462.D	Decachlorobiphenyl	1	20	19.7	98		43	140
		Tetrachloro-m-xylene	1	20	20.9	104		77	126
		Decachlorobiphenyl	2	20	18.3	91		43	140
		Tetrachloro-m-xylene	2	20	20.9	105		77	126