

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: ROYF02
 Lab Code: CHEM Case No.: Q1421 SAS No.: Q1421 SDG No.: Q1421
 Instrument ID: MSVOA_N Calibration Date(s): 02/18/2025 02/18/2025
 Heated Purge: (Y/N) N Calibration Time(s): 11:09 14:18
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF100 = VN085772.D		RRF050 = VN085773.D		RRF010 = VN085775.D		
		RRF005 = VN085776.D		RRF001 = VN085777.D		RRF020 = VN085779.D		
COMPOUND	RRF100	RRF050	RRF010	RRF005	RRF001	RRF020	RRF	% RSD
Dichlorodifluoromethane	0.694	0.669	0.779	0.650	0.660	0.628	0.680	7.8
Chloromethane	0.647	0.622	0.742	0.662	0.817	0.604	0.682	11.9
Vinyl Chloride	0.698	0.678	0.796	0.729	0.761	0.674	0.723	6.7
Bromomethane	0.434	0.424	0.543	0.496		0.413	0.462	12.1
Chloroethane	0.447	0.417	0.516	0.470	0.584	0.438	0.479	12.9
Trichlorofluoromethane	1.010	0.959	1.134	1.065	1.103	1.006	1.046	6.3
1,1,2-Trichlorotrifluoroethane	0.602	0.564	0.674	0.635	0.554	0.599	0.605	7.4
1,1-Dichloroethene	0.555	0.527	0.591	0.571	0.514	0.531	0.548	5.3
Acetone	0.195	0.184	0.217	0.208	0.227	0.198	0.205	7.7
Carbon Disulfide	1.565	1.473	1.804	1.604	1.848	1.455	1.625	10.2
Methyl tert-butyl Ether	1.747	1.658	1.714	1.560	1.438	1.659	1.629	7
Methyl Acetate	0.653	0.619	0.794	0.690	0.853	0.625	0.705	13.7
Methylene Chloride	0.628	0.593	0.710	0.666	0.736	0.625	0.660	8.3
trans-1,2-Dichloroethene	0.585	0.549	0.621	0.599	0.623	0.566	0.590	5.1
1,1-Dichloroethane	1.097	1.035	1.216	1.125	1.073	1.089	1.106	5.6
Cyclohexane	0.878	0.860	0.977	1.056		0.858	0.926	9.5
2-Butanone	0.308	0.294	0.331	0.299	0.279	0.299	0.302	5.6
Carbon Tetrachloride	0.573	0.528	0.583	0.545	0.533	0.538	0.550	4.1
cis-1,2-Dichloroethene	0.705	0.661	0.740	0.685	0.646	0.671	0.685	4.9
Bromochloromethane	0.481	0.444	0.346	0.570	0.483	0.402	0.454	17
Chloroform	1.120	1.070	1.269	1.164	1.171	1.141	1.156	5.7
1,1,1-Trichloroethane	1.012	0.967	1.089	1.058	1.030	1.012	1.028	4.1
Methylcyclohexane	0.560	0.492	0.446	0.405	0.373	0.453	0.455	14.5
Benzene	1.581	1.441	1.568	1.421	1.420	1.474	1.484	4.9
1,2-Dichloroethane	0.499	0.456	0.527	0.464	0.472	0.483	0.483	5.4
Trichloroethene	0.370	0.339	0.384	0.350	0.395	0.348	0.364	6.1
1,2-Dichloropropane	0.377	0.347	0.385	0.344	0.339	0.357	0.358	5.2
Bromodichloromethane	0.569	0.518	0.571	0.533	0.512	0.541	0.541	4.6
4-Methyl-2-Pentanone	0.414	0.379	0.392	0.342	0.290	0.388	0.367	12.1
Toluene	0.993	0.902	0.924	0.820	0.689	0.891	0.870	12

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: ROYF02
 Lab Code: CHEM Case No.: Q1421 SAS No.: Q1421 SDG No.: Q1421
 Instrument ID: MSVOA_N Calibration Date(s): 02/18/2025 02/18/2025
 Heated Purge: (Y/N) N Calibration Time(s): 11:09 14:18
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF100 = VN085772.D		RRF050 = VN085773.D		RRF010 = VN085775.D			
		RRF005 = VN085776.D		RRF001 = VN085777.D		RRF020 = VN085779.D			
COMPOUND	RRF100	RRF050	RRF010	RRF005	RRF001	RRF020	RRF	% RSD	
t-1,3-Dichloropropene	0.576	0.521	0.518	0.471	0.408	0.505	0.500	11.3	
cis-1,3-Dichloropropene	0.626	0.566	0.582	0.513	0.442	0.559	0.548	11.6	
1,1,2-Trichloroethane	0.361	0.331	0.370	0.346	0.324	0.344	0.346	5	
2-Hexanone	0.296	0.273	0.266	0.227	0.184	0.270	0.253	16	
Dibromochloromethane	0.440	0.402	0.431	0.384	0.355	0.405	0.403	7.7	
1,2-Dibromoethane	0.354	0.322	0.362	0.320	0.273	0.330	0.327	9.7	
Tetrachloroethene	0.384	0.361	0.420	0.391	0.393	0.375	0.387	5.2	
Chlorobenzene	1.183	1.110	1.212	1.139	1.109	1.119	1.145	3.8	
Ethyl Benzene	2.105	1.904	1.830	1.665	1.424	1.838	1.794	12.8	
m/p-Xylenes	0.809	0.741	0.725	0.610	0.481	0.722	0.682	17.2	
o-Xylene	0.771	0.704	0.666	0.584	0.502	0.659	0.648	14.5	
Styrene	1.327	1.203	1.060	0.885	0.744	1.133	1.059	20.2	
Bromoform	0.327	0.303	0.322	0.301	0.274	0.309	0.306	6.2	
Isopropylbenzene	3.851	3.596	3.623	3.067	2.933	3.484	3.425	10.3	
1,1,2,2-Tetrachloroethane	1.073	1.053	1.304	1.218	1.271	1.154	1.179	8.8	
1,3-Dichlorobenzene	1.772	1.635	1.876	1.697	1.748	1.675	1.734	4.9	
1,4-Dichlorobenzene	1.750	1.647	1.884	1.777	1.921	1.684	1.777	6.1	
1,2-Dichlorobenzene	1.660	1.574	1.768	1.642	1.744	1.617	1.667	4.5	
1,2-Dibromo-3-Chloropropane	0.198	0.193	0.238	0.222	0.256	0.191	0.216	12.4	
1,2,4-Trichlorobenzene	0.877	0.802	0.827	0.728	0.829	0.704	0.794	8.3	
1,2,3-Trichlorobenzene	0.825	0.789	0.844	0.736	0.917	0.675	0.798	10.6	
1,2-Dichloroethane-d4	0.698	0.613	0.588	0.767		0.575	0.648	12.6	
Dibromofluoromethane	0.375	0.317	0.294	0.357		0.293	0.327	11.5	
Toluene-d8	1.444	1.199	1.017	1.212		1.080	1.190	13.8	
4-Bromofluorobenzene	0.503	0.419	0.323	0.370		0.346	0.392	18.2	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.