

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ROYF02
 Lab Code: CHEM Case No.: Q1421 SAS No.: Q1421 SDG NO.: Q1421
 Lab File ID: VN085862.D Date Analyzed: 02/26/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:32
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	218212	13.788				
UPPER LIMIT	436424	14.288				
LOWER LIMIT	109106	13.288				
EPA SAMPLE NO.						
P001-CLAY-CF01-01	167346	13.79				
P001-CLAY-CF01-01MS	197597	13.79				
P001-CLAY-CF01-01MSD	196698	13.79				
P001-CLAY-CF01-02	165447	13.79				
P001-CLAY-CF02-01	172512	13.79				
VN0226WBL01	165756	13.79				
VN0226WBS01	219511	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.