

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ROYF02

Lab Code: CHEM Case No.: Q1421 SAS No.: Q1421 SDG No.: Q1421

Instrument ID: MSVOA_Y Calibration Date/Time: 02/28/2025 09:44

Lab File ID: VY021375.D Init. Calib. Date(s): 02/25/2025 02/25/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 12:40 16:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.474	0.440		-7.17	20
Chloromethane	0.612	0.566	0.1	-7.52	20
Vinyl Chloride	0.601	0.616		2.5	20
Bromomethane	0.399	0.411		3.01	20
Chloroethane	0.373	0.400		7.24	20
Trichlorofluoromethane	0.908	0.936		3.08	20
1,1,2-Trichlorotrifluoroethane	0.525	0.511		-2.67	20
1,1-Dichloroethene	0.503	0.487		-3.18	20
Acetone	0.109	0.132		21.1	20
Carbon Disulfide	1.671	1.585		-5.15	20
Methyl tert-butyl Ether	1.329	1.304		-1.88	20
Methyl Acetate	0.266	0.247		-7.14	20
Methylene Chloride	0.588	0.547		-6.97	20
trans-1,2-Dichloroethene	0.554	0.549		-0.9	20
1,1-Dichloroethane	1.040	1.009	0.1	-2.98	20
Cyclohexane	1.000	0.868		-13.2	20
2-Butanone	0.164	0.166		1.22	20
Carbon Tetrachloride	0.554	0.578		4.33	20
cis-1,2-Dichloroethene	0.632	0.635		0.47	20
Bromochloromethane	0.460	0.462		0.44	20
Chloroform	1.054	1.043		-1.04	20
1,1,1-Trichloroethane	0.960	0.951		-0.94	20
Methylcyclohexane	0.623	0.607		-2.57	20
Benzene	1.456	1.487		2.13	20
1,2-Dichloroethane	0.413	0.436		5.57	20
Trichloroethene	0.364	0.375		3.02	20
1,2-Dichloropropane	0.350	0.356		1.71	20
Bromodichloromethane	0.511	0.537		5.09	20
4-Methyl-2-Pentanone	0.243	0.242		-0.41	20
Toluene	0.904	0.955		5.64	20
t-1,3-Dichloropropene	0.468	0.488		4.27	20
cis-1,3-Dichloropropene	0.552	0.565		2.36	20
1,1,2-Trichloroethane	0.249	0.254		2.01	20
2-Hexanone	0.159	0.168		5.66	20
Dibromochloromethane	0.340	0.362		6.47	20
1,2-Dibromoethane	0.236	0.241		2.12	20
Tetrachloroethene	0.378	0.394		4.23	20
Chlorobenzene	1.122	1.154	0.3	2.85	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ROYF02

Lab Code: CHEM Case No.: Q1421 SAS No.: Q1421 SDG No.: Q1421

Instrument ID: MSVOA_Y Calibration Date/Time: 02/28/2025 09:44

Lab File ID: VY021375.D Init. Calib. Date(s): 02/25/2025 02/25/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 12:40 16:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.988	2.041		2.67	20
m/p-Xylenes	0.748	0.775		3.61	20
o-Xylene	0.701	0.722		3	20
Styrene	1.166	1.235		5.92	20
Bromoform	0.225	0.232	0.1	3.11	20
Isopropylbenzene	3.759	3.861		2.71	20
1,1,2,2-Tetrachloroethane	0.663	0.654	0.3	-1.36	20
1,3-Dichlorobenzene	1.730	1.775		2.6	20
1,4-Dichlorobenzene	1.709	1.709		0	20
1,2-Dichlorobenzene	1.515	1.525		0.66	20
1,2-Dibromo-3-Chloropropane	0.105	0.101		-3.81	20
1,2,4-Trichlorobenzene	0.925	0.898		-2.92	20
1,2,3-Trichlorobenzene	0.781	0.738		-5.51	20
1,2-Dichloroethane-d4	0.567	0.567		0	20
Dibromofluoromethane	0.328	0.351		7.01	20
Toluene-d8	1.267	1.327		4.74	20
4-Bromofluorobenzene	0.426	0.441		3.52	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.