

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY022525

Review By	Mahesh Dadoda	Review On	2/26/2025 1:45:35 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	2/26/2025 1:46:55 PM		
SubDirectory	VY022525	HP Acquire Method	MSVOA_Y	HP Processing Method	82y022525s.m

STD. NAME	STD REF.#
Tune/Reschk	VP133154
Initial Calibration Stds	VP133137,VP133140,VP133143,VP133146,VP133147,VP133149
CCC	
Internal Standard/PEM	VP131783
ICV/I.BLK	VP133150
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021308.D	25 Feb 2025 12:10		SY/MD	Ok
2	VSTDIC005	VSTDIC005	VY021309.D	25 Feb 2025 12:40		SY/MD	Ok,M
3	VSTDIC010	VSTDIC010	VY021310.D	25 Feb 2025 13:03		SY/MD	Ok,M
4	VSTDIC020	VSTDIC020	VY021311.D	25 Feb 2025 13:26	Acrolein fail for method	SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY021312.D	25 Feb 2025 14:48		SY/MD	Ok,M
6	VSTDIC100	VSTDIC100	VY021313.D	25 Feb 2025 15:34	Not used	SY/MD	Not Ok
7	VSTDIC150	VSTDIC150	VY021314.D	25 Feb 2025 15:57		SY/MD	Ok,M
8	VIBLK	VIBLK	VY021315.D	25 Feb 2025 16:24		SY/MD	Ok
9	VSTDIC100	VSTDIC100	VY021316.D	25 Feb 2025 16:49		SY/MD	Ok,M
10	VSTDICV050	ICVVY022525	VY021317.D	25 Feb 2025 17:31	ICV Failed for comp. #10	SY/MD	Ok,M
11	VIBLK	VIBLK	VY021318.D	25 Feb 2025 17:54		SY/MD	Ok

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY022725

Review By	Mahesh Dadoda	Review On	2/28/2025 11:07:46 AM
Supervise By	Semsettin Yesilyurt	Supervise On	2/28/2025 11:11:02 AM
SubDirectory	VY022725	HP Acquire Method	HP Processing Method 82y022225s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133167 VP133165,VP133166

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021345.D	27 Feb 2025 09:41		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021346.D	27 Feb 2025 10:13		SY/MD	Ok,M
3	VY0227SBL01	VY0227SBL01	VY021347.D	27 Feb 2025 10:42		SY/MD	Ok
4	VY0227SBS01	VY0227SBS01	VY021348.D	27 Feb 2025 11:18		SY/MD	Ok,M
5	VY0227SBSD01	VY0227SBSD01	VY021349.D	27 Feb 2025 11:40		SY/MD	Ok,M
6	Q1428-01	NB-07-022525	VY021350.D	27 Feb 2025 12:04	Vial-A ISTD Fail	SY/MD	ReRun
7	Q1442-03	351	VY021351.D	27 Feb 2025 12:28	Vial-A	SY/MD	Ok
8	Q1416-01RE	OK-01-022425RE	VY021352.D	27 Feb 2025 12:51	Vial-B ISTD Fail	SY/MD	Confirms
9	Q1428-01RE	NB-07-022525RE	VY021353.D	27 Feb 2025 13:14	Vial-B Internal Standard Fail	SY/MD	Confirms
10	Q1418-01	TR-06-02242025	VY021354.D	27 Feb 2025 13:38	Vial-B Not purge	SY/MD	Not Ok
11	Q1415-03RE	B-163-SB02RE	VY021355.D	27 Feb 2025 14:01	Vial-B Internal standard fail;Not match for com.#20	SY/MD	Confirms
12	Q1422-02	B-154-SB02	VY021356.D	27 Feb 2025 14:25	Vial-A Internal standard fail	SY/MD	ReRun
13	Q1422-01	B-154-SB01	VY021357.D	27 Feb 2025 14:48	Vial-A Internal standard fail	SY/MD	ReRun
14	Q1421-09	P001-CLAY-CF02-01	VY021358.D	27 Feb 2025 15:12	Internal Standard Fail	SY/MD	ReRun
15	Q1421-07	P001-CLAY-CF01-02	VY021359.D	27 Feb 2025 15:35	Vial-A Internal standard fail	SY/MD	ReRun
16	Q1421-01	P001-CLAY-CF01-01	VY021360.D	27 Feb 2025 15:58	Vial-A Internal standard fail	SY/MD	Ok
17	Q1427-01RE	VNJ-227RE	VY021361.D	27 Feb 2025 16:22	Vial-B Internal standard fail;Surrogate fail	SY/MD	Confirms

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY022725

Review By	Maresh Dadoda	Review On	2/28/2025 11:07:46 AM
Supervise By	Semsettin Yesilyurt	Supervise On	2/28/2025 11:11:02 AM
SubDirectory	VY022725	HP Acquire Method	HP Processing Method 82y022225s.m
STD. NAME	STD REF.#		
Tune/Reschk	VP133167		
Initial Calibration Stds	VP133165,VP133166		
CCC			
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	Q1420-07	TP-2-WC-VOC	VY021362.D	27 Feb 2025 16:45	Vial-B	SY/MD	Ok
19	Q1411-03	HR-03-02212025	VY021363.D	27 Feb 2025 17:09	Vial-B Not purge	SY/MD	Not Ok
20	Q1411-01	HR-02-02212025	VY021364.D	27 Feb 2025 17:32	Vial-B Internal standard fail	SY/MD	Ok
21	Q1440-01	OR-02-022625	VY021365.D	27 Feb 2025 17:55	Vial-A Internal standard fail	SY/MD	ReRun
22	Q1448-01	PSP1	VY021366.D	27 Feb 2025 18:19	Vial-A Internal standard fail	SY/MD	ReRun
23	Q1448-02	P1	VY021367.D	27 Feb 2025 18:42	Vial-A Internal standard fail	SY/MD	ReRun
24	Q1448-03	P2	VY021368.D	27 Feb 2025 19:06	Vial-A Internal standard fail	SY/MD	ReRun
25	Q1448-04	P3	VY021369.D	27 Feb 2025 19:29	Vial-A Internal standard fail;Need MeOH	SY/MD	Dilution
26	Q1448-05	P4	VY021370.D	27 Feb 2025 19:52	Vial-A Internal standard fail	SY/MD	ReRun
27	Q1421-02MS	P001-CLAY-CF01-01M	VY021371.D	27 Feb 2025 20:15	Vial-A Internal standard fail	SY/MD	Ok,M
28	Q1421-03MSD	P001-CLAY-CF01-01M	VY021372.D	27 Feb 2025 20:38	Vial-A Internal standard fail	SY/MD	Ok,M
29	VSTDCCC050	VSTDCCC050EC	VY021373.D	27 Feb 2025 21:01		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY022825

Review By	John Carlone	Review On	3/3/2025 12:12:17 PM
Supervise By	Maresh Dadoda	Supervise On	3/3/2025 1:51:10 PM
SubDirectory	VY022825	HP Acquire Method	MSVOA_Y HP Processing Method 82y022525s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133182
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133183,VP133184 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021374.D	28 Feb 2025 09:03		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021375.D	28 Feb 2025 09:44	CCAL failed low for comp. #11; CCAL failed high for comp. #16	SY/MD	Ok,M
3	VY0228SBL01	VY0228SBL01	VY021376.D	28 Feb 2025 10:59		SY/MD	Ok
4	VY0228SBS01	VY0228SBS01	VY021377.D	28 Feb 2025 11:32		SY/MD	Ok,M
5	VY0228SBSD01	VY0228SBSD01	VY021378.D	28 Feb 2025 11:55		SY/MD	Ok,M
6	Q1421-07RE	P001-CLAY-CF01-02RE	VY021379.D	28 Feb 2025 12:56	Vial-B Internal Standard Fail	SY/MD	Confirms
7	Q1422-01RE	B-154-SB01RE	VY021380.D	28 Feb 2025 13:20	Vial-B Internal Standard Fail	SY/MD	Confirms
8	Q1422-02	B-154-SB02	VY021381.D	28 Feb 2025 15:31	Vial-B	SY/MD	Ok
9	Q1448-01RE	PSP1RE	VY021382.D	28 Feb 2025 15:55	Vial-B CCAL failed low for comp. #11	SY/MD	Confirms
10	Q1448-02RE	P1RE	VY021383.D	28 Feb 2025 16:18	Vial-B CCAL failed low for comp. #11	SY/MD	Confirms
11	Q1448-03RE	P2RE	VY021384.D	28 Feb 2025 16:41	Vial-B CCAL failed low for comp. #11;Not match for comp. #20	SY/MD	Confirms
12	Q1448-05RE	P4RE	VY021385.D	28 Feb 2025 17:05	Vial-B CCAL failed low for comp. #11; CCAL failed high for comp. #16;Not match for comp. #11,16,20	SY/MD	Confirms
13	Q1440-01	OR-02-022625	VY021386.D	28 Feb 2025 17:28	Vial-B	SY/MD	Ok
14	Q1441-01	HD-02-022625	VY021387.D	28 Feb 2025 17:52	Vial-B Internal Standard Fail	SY/MD	ReRun
15	Q1458-02	SP-SOIL-VOC	VY021388.D	28 Feb 2025 18:15	vial-A	SY/MD	Ok,M

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY022825

Review By	John Carlone	Review On	3/3/2025 12:12:17 PM
Supervise By	Maresh Dadoda	Supervise On	3/3/2025 1:51:10 PM
SubDirectory	VY022825	HP Acquire Method	MSVOA_Y HP Processing Method 82y022525s.m
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133182		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133183,VP133184 VP131783		

16	Q1463-01	T1	VY021389.D	28 Feb 2025 18:38	Vial-A CCAL failed low for comp. #11	SY/MD	ReRun
17	Q1457-01	HR-0-02272025	VY021390.D	28 Feb 2025 19:02	Vial-A	SY/MD	Ok
18	Q1463-03	T3	VY021391.D	28 Feb 2025 19:25	Vial-A Internal Standard Fail; CCAL failed low for comp. #11; CCAL failed high for comp.#16	SY/MD	ReRun
19	Q1463-04	T4	VY021392.D	28 Feb 2025 19:49	Vial-A CCAL failed low for comp. #11	SY/MD	ReRun
20	Q1463-05	T5	VY021393.D	28 Feb 2025 20:12	Vial-A Internal Standard Fail; CCAL failed low for comp. #11; CCAL failed high for comp.#16	SY/MD	ReRun
21	VSTDCCC050	VSTDCCC050EC	VY021394.D	28 Feb 2025 20:35		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY030425

Review By	Mahesh Dadoda	Review On	3/5/2025 12:14:39 PM
Supervise By	Semsettin Yesilyurt	Supervise On	3/5/2025 12:18:08 PM
SubDirectory	VY030425	HP Acquire Method	MSVOA_Y HP Processing Method 82y030425s.m

STD. NAME	STD REF.#
Tune/Reschk	VP133216
Initial Calibration Stds	VP133207,VP133208,VP133209,VP133210,VP133211,VP133212
CCC	VP133214,VP133215
Internal Standard/PEM	VP131783
ICV/I.BLK	VP133213
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021395.D	04 Mar 2025 08:52		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY021396.D	04 Mar 2025 09:23	not used	SY/MD	Not Ok
3	VSTDICC010	VSTDICC010	VY021397.D	04 Mar 2025 09:46		SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY021398.D	04 Mar 2025 10:09	Comp.#11 is on Linear Regression	SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY021399.D	04 Mar 2025 10:30	Comp.#95 is on Quadratic Regression	SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY021400.D	04 Mar 2025 11:06		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY021401.D	04 Mar 2025 11:29	Method fail for comp.#13	SY/MD	Ok,M
8	VIBLK	VIBLK	VY021402.D	04 Mar 2025 11:52		SY/MD	Ok
9	VSTDICC005	VSTDICC005	VY021403.D	04 Mar 2025 12:15		SY/MD	Ok,M
10	VSTDICV050	ICVVY030425	VY021404.D	04 Mar 2025 13:12		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY030625

Review By	Maresh Dadoda	Review On	3/7/2025 1:32:16 PM
Supervise By	Semsettin Yesilyurt	Supervise On	3/7/2025 1:32:53 PM
SubDirectory	VY030625	HP Acquire Method	HP Processing Method 82y030425s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133223
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133224,VP133225 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021425.D	06 Mar 2025 09:39		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021426.D	06 Mar 2025 10:10		SY/MD	Ok,M
3	VY0306SBL01	VY0306SBL01	VY021427.D	06 Mar 2025 11:59		SY/MD	Ok
4	VY0306SBS01	VY0306SBS01	VY021428.D	06 Mar 2025 12:33		SY/MD	Ok,M
5	VY0306SBSD01	VY0306SBSD01	VY021429.D	06 Mar 2025 12:55		SY/MD	Ok,M
6	Q1489-05	TP-2	VY021430.D	06 Mar 2025 13:28	vial-B	SY/MD	Ok
7	Q1489-01	TP-1	VY021431.D	06 Mar 2025 13:52	vial-B Not purge	SY/MD	Not Ok
8	Q1421-09	P001-CLAY-CF02-01	VY021432.D	06 Mar 2025 14:15	vial-B	SY/MD	Ok
9	Q1479-01	P5	VY021433.D	06 Mar 2025 14:39	vial-A	SY/MD	Ok
10	Q1479-02	DSP2	VY021434.D	06 Mar 2025 15:02	vial-A Internal Standard Fail	SY/MD	ReRun
11	Q1488-01	ENV-101-SB01	VY021435.D	06 Mar 2025 15:25	vial-A	SY/MD	Ok
12	Q1488-03	ENV-101-SB02	VY021436.D	06 Mar 2025 15:49	vial-A	SY/MD	Ok
13	Q1488-05	ENV-102-SB01	VY021437.D	06 Mar 2025 16:12	vial-A	SY/MD	Ok
14	Q1488-07	ENV-102-SB02	VY021438.D	06 Mar 2025 16:36	vial-A Not purge	SY/MD	Not Ok
15	Q1488-09	ENV-104-SB01	VY021439.D	06 Mar 2025 16:59	vial-A Internal Standard Fail	SY/MD	ReRun
16	Q1488-11	ENV-104-SB02	VY021440.D	06 Mar 2025 17:22	vial-A Not purge	SY/MD	Not Ok
17	Q1463-01	T1	VY021441.D	06 Mar 2025 17:46	vial-B	SY/MD	Ok
18	Q1463-05	T5	VY021442.D	06 Mar 2025 18:09	vial-B	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY030625

Review By	Maresh Dadoda	Review On	3/7/2025 1:32:16 PM
Supervise By	Semsettin Yesilyurt	Supervise On	3/7/2025 1:32:53 PM
SubDirectory	VY030625	HP Acquire Method	HP Processing Method 82y030425s.m
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133223		
CCC	VP133224,VP133225		
Internal Standard/PEM	VP131783		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q1463-04	T4	VY021443.D	06 Mar 2025 18:33	Vial-B	SY/MD	Ok
20	Q1463-06	T6	VY021444.D	06 Mar 2025 18:56	Vial-A	SY/MD	Ok
21	Q1463-07	T7	VY021445.D	06 Mar 2025 19:20	Vial-A	SY/MD	Ok
22	Q1463-08	T8	VY021446.D	06 Mar 2025 19:43	Vial-A	SY/MD	Ok
23	Q1463-09	T9	VY021447.D	06 Mar 2025 20:06	Vial-A	SY/MD	Ok
24	Q1463-10	T10	VY021448.D	06 Mar 2025 20:30	Vial-A Internal Standard Fail; Surrogate Fail	SY/MD	ReRun
25	VSTDCCC050	VSTDCCC050EC	VY021449.D	06 Mar 2025 21:15		SY/MD	Ok,M

M : Manual Integration