

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1421

Client: Weston Solutions, Inc.

Analytical Method: SW8260D

Datafile : VY021377.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0228SBS01	Dichlorodifluoromethane	20	18.6	ug/Kg	93			64	136	
	Chloromethane	20	17.7	ug/Kg	89			70	130	
	Vinyl chloride	20	19.3	ug/Kg	97			72	129	
	Bromomethane	20	20.1	ug/Kg	101			58	141	
	Chloroethane	20	20.1	ug/Kg	101			69	130	
	Trichlorofluoromethane	20	20.0	ug/Kg	100			69	134	
	1,1,2-Trichlorotrifluoroethane	20	19.7	ug/Kg	99			81	123	
	1,1-Dichloroethene	20	18.8	ug/Kg	94			79	121	
	Acetone	100	77.6	ug/Kg	78			60	131	
	Carbon disulfide	20	17.4	ug/Kg	87			45	154	
	Methyl tert-butyl Ether	20	18.8	ug/Kg	94			77	129	
	Methyl Acetate	20	18.4	ug/Kg	92			69	149	
	Methylene Chloride	20	19.1	ug/Kg	96			56	174	
	trans-1,2-Dichloroethene	20	18.4	ug/Kg	92			80	123	
	1,1-Dichloroethane	20	18.2	ug/Kg	91			82	123	
	Cyclohexane	20	17.1	ug/Kg	86			76	122	
	2-Butanone	100	83.5	ug/Kg	84			69	131	
	Carbon Tetrachloride	20	19.5	ug/Kg	98			76	129	
	cis-1,2-Dichloroethene	20	18.7	ug/Kg	94			82	123	
	Bromochloromethane	20	17.8	ug/Kg	89			80	127	
	Chloroform	20	18.9	ug/Kg	95			82	125	
	1,1,1-Trichloroethane	20	19.1	ug/Kg	96			80	126	
	Methylcyclohexane	20	18.1	ug/Kg	91			77	123	
	Benzene	20	19.1	ug/Kg	96			84	121	
	1,2-Dichloroethane	20	19.5	ug/Kg	98			81	126	
	Trichloroethene	20	19.6	ug/Kg	98			83	122	
	1,2-Dichloropropane	20	19.0	ug/Kg	95			83	122	
	Bromodichloromethane	20	19.7	ug/Kg	99			82	123	
	4-Methyl-2-Pentanone	100	92.3	ug/Kg	92			70	135	
	Toluene	20	19.3	ug/Kg	97			83	122	
	t-1,3-Dichloropropene	20	18.6	ug/Kg	93			78	124	
	cis-1,3-Dichloropropene	20	18.7	ug/Kg	94			81	122	
	1,1,2-Trichloroethane	20	19.7	ug/Kg	99			82	125	
	2-Hexanone	100	91.3	ug/Kg	91			66	138	
	Dibromochloromethane	20	20.4	ug/Kg	102			79	125	
	1,2-Dibromoethane	20	19.5	ug/Kg	98			80	125	
	Tetrachloroethene	20	20.4	ug/Kg	102			83	125	
	Chlorobenzene	20	19.7	ug/Kg	99			84	122	
	Ethyl Benzene	20	19.1	ug/Kg	96			82	124	
	m/p-Xylenes	40	39.2	ug/Kg	98			83	124	
	o-Xylene	20	19.4	ug/Kg	97			83	123	
	Styrene	20	19.6	ug/Kg	98			82	124	
	Bromoform	20	20.2	ug/Kg	101			75	127	
	Isopropylbenzene	20	18.9	ug/Kg	95			82	124	
	1,1,2,2-Tetrachloroethane	20	19.2	ug/Kg	96			77	127	
	1,3-Dichlorobenzene	20	19.5	ug/Kg	98			83	122	
	1,4-Dichlorobenzene	20	19.3	ug/Kg	97			84	121	
	1,2-Dichlorobenzene	20	19.5	ug/Kg	98			83	124	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1421

Client: Weston Solutions, Inc.

Analytical Method: SW8260D

Datafile : VY021377.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0228SBS01	1,2-Dibromo-3-Chloropropane	20	18.5	ug/Kg	93			66	134	
	1,2,4-Trichlorobenzene	20	18.4	ug/Kg	92			78	127	
	1,2,3-Trichlorobenzene	20	18.7	ug/Kg	94			70	137	