

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: ROYF02
Lab Code: CHEM Case No.: Q1421 SAS No.: Q1421 SDG NO.: Q1421
Lab File ID: VY021425.D BFB Injection Date: 03/06/2025
Instrument ID: MSVOA_Y BFB Injection Time: 09:39
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	21.2
75	30.0 - 60.0% of mass 95	54
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.9 (1.1) 1
174	50.0 - 100.0% of mass 95	82.3
175	5.0 - 9.0% of mass 174	6.2 (7.5) 1
176	95.0 - 101.0% of mass 174	79.6 (96.8) 1
177	5.0 - 9.0% of mass 176	5.7 (7.1) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY021426.D	03/06/2025	10:10
VY0306SBL01	VY0306SBL01	VY021427.D	03/06/2025	11:59
VY0306SBS01	VY0306SBS01	VY021428.D	03/06/2025	12:33
P001-CLAY-CF02-01	Q1421-09	VY021432.D	03/06/2025	14:15