

Initial and Continuing Calibration Verification

Client: Portal Partners Tri-Venture
Project: Amtrak Sawtooth Bridges 2025

SDG No.: Q1422
RunNo.: LB134815

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.501	0.5	100	90-110	02/26/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.497	0.5	99	90-110	02/26/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.499	0.5	100	90-110	02/26/2025
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.499	0.5	100	90-110	02/26/2025