

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q1422
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1422 **SAS No.:** Q1422
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV19	Mercury	3.63	4.0	91	90 - 110	CV	02/26/2025	11:10	LB134807

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q1422
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1422 **SAS No.:** Q1422
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV25	Mercury	5.01	5.0	100	90 - 110	CV	02/26/2025	11:18	LB134807
CCV26	Mercury	4.86	5.0	97	90 - 110	CV	02/26/2025	11:48	LB134807
CCV27	Mercury	4.78	5.0	96	90 - 110	CV	02/26/2025	12:30	LB134807
CCV28	Mercury	4.73	5.0	95	90 - 110	CV	02/26/2025	12:46	LB134807

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture SDG No.: Q1422
Contract: PORT06 Lab Code: CHEM Case No.: Q1422 SAS No.: Q1422
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	2500	2500	100	90 - 110	P	03/10/2025	11:48	LB134965
	Antimony	1010	1000	100	90 - 110	P	03/10/2025	11:48	LB134965
	Arsenic	1030	1000	103	90 - 110	P	03/10/2025	11:48	LB134965
	Barium	537	520	103	90 - 110	P	03/10/2025	11:48	LB134965
	Beryllium	517	510	101	90 - 110	P	03/10/2025	11:48	LB134965
	Cadmium	503	510	99	90 - 110	P	03/10/2025	11:48	LB134965
	Calcium	9910	10000	99	90 - 110	P	03/10/2025	11:48	LB134965
	Chromium	531	520	102	90 - 110	P	03/10/2025	11:48	LB134965
	Cobalt	510	520	98	90 - 110	P	03/10/2025	11:48	LB134965
	Copper	524	510	103	90 - 110	P	03/10/2025	11:48	LB134965
	Iron	10400	10000	104	90 - 110	P	03/10/2025	11:48	LB134965
	Lead	1010	1000	101	90 - 110	P	03/10/2025	11:48	LB134965
	Magnesium	5800	6000	97	90 - 110	P	03/10/2025	11:48	LB134965
	Manganese	503	520	97	90 - 110	P	03/10/2025	11:48	LB134965
	Nickel	513	530	97	90 - 110	P	03/10/2025	11:48	LB134965
	Potassium	10200	9900	103	90 - 110	P	03/10/2025	11:48	LB134965
	Selenium	1040	1000	104	90 - 110	P	03/10/2025	11:48	LB134965
	Silver	257	250	103	90 - 110	P	03/10/2025	11:48	LB134965
	Sodium	9760	10000	98	90 - 110	P	03/10/2025	11:48	LB134965
	Thallium	1040	1000	104	90 - 110	P	03/10/2025	11:48	LB134965
	Vanadium	513	500	103	90 - 110	P	03/10/2025	11:48	LB134965
	Zinc	1040	1000	104	90 - 110	P	03/10/2025	11:48	LB134965

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q1422

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1422 **SAS No.:** Q1422

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	103	100	103	80 - 120	P	03/10/2025	11:52	LB134965
	Antimony	49.7	50.0	99	80 - 120	P	03/10/2025	11:52	LB134965
	Arsenic	17.9	20.0	90	80 - 120	P	03/10/2025	11:52	LB134965
	Barium	90.3	100	90	80 - 120	P	03/10/2025	11:52	LB134965
	Beryllium	5.54	6.0	92	80 - 120	P	03/10/2025	11:52	LB134965
	Cadmium	5.56	6.0	93	80 - 120	P	03/10/2025	11:52	LB134965
	Calcium	1900	2000	95	80 - 120	P	03/10/2025	11:52	LB134965
	Chromium	9.58	10.0	96	80 - 120	P	03/10/2025	11:52	LB134965
	Cobalt	28.0	30.0	93	80 - 120	P	03/10/2025	11:52	LB134965
	Copper	21.2	20.0	106	80 - 120	P	03/10/2025	11:52	LB134965
	Iron	108	100	108	80 - 120	P	03/10/2025	11:52	LB134965
	Lead	11.7	12.0	97	80 - 120	P	03/10/2025	11:52	LB134965
	Magnesium	1970	2000	99	80 - 120	P	03/10/2025	11:52	LB134965
	Manganese	19.2	20.0	96	80 - 120	P	03/10/2025	11:52	LB134965
	Nickel	37.4	40.0	93	80 - 120	P	03/10/2025	11:52	LB134965
	Potassium	1930	2000	97	80 - 120	P	03/10/2025	11:52	LB134965
	Selenium	18.2	20.0	91	80 - 120	P	03/10/2025	11:52	LB134965
	Silver	9.78	10.0	98	80 - 120	P	03/10/2025	11:52	LB134965
	Sodium	1800	2000	90	80 - 120	P	03/10/2025	11:52	LB134965
	Thallium	40.0	40.0	100	80 - 120	P	03/10/2025	11:52	LB134965
	Vanadium	35.7	40.0	89	80 - 120	P	03/10/2025	11:52	LB134965
	Zinc	40.5	40.0	101	80 - 120	P	03/10/2025	11:52	LB134965

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q1422

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1422 **SAS No.:** Q1422

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9760	10000	98	90 - 110	P	03/10/2025	12:27	LB134965
	Antimony	5060	5000	101	90 - 110	P	03/10/2025	12:27	LB134965
	Arsenic	4990	5000	100	90 - 110	P	03/10/2025	12:27	LB134965
	Barium	9500	10000	95	90 - 110	P	03/10/2025	12:27	LB134965
	Beryllium	247	250	99	90 - 110	P	03/10/2025	12:27	LB134965
	Cadmium	2440	2500	97	90 - 110	P	03/10/2025	12:27	LB134965
	Calcium	26100	25000	104	90 - 110	P	03/10/2025	12:27	LB134965
	Chromium	1000	1000	100	90 - 110	P	03/10/2025	12:27	LB134965
	Cobalt	2430	2500	97	90 - 110	P	03/10/2025	12:27	LB134965
	Copper	1240	1250	99	90 - 110	P	03/10/2025	12:27	LB134965
	Iron	4820	5000	96	90 - 110	P	03/10/2025	12:27	LB134965
	Lead	4900	5000	98	90 - 110	P	03/10/2025	12:27	LB134965
	Magnesium	25900	25000	104	90 - 110	P	03/10/2025	12:27	LB134965
	Manganese	2620	2500	105	90 - 110	P	03/10/2025	12:27	LB134965
	Nickel	2440	2500	98	90 - 110	P	03/10/2025	12:27	LB134965
	Potassium	24200	25000	97	90 - 110	P	03/10/2025	12:27	LB134965
	Selenium	5030	5000	101	90 - 110	P	03/10/2025	12:27	LB134965
	Silver	1220	1250	98	90 - 110	P	03/10/2025	12:27	LB134965
	Sodium	24100	25000	96	90 - 110	P	03/10/2025	12:27	LB134965
	Thallium	4960	5000	99	90 - 110	P	03/10/2025	12:27	LB134965
	Vanadium	2410	2500	96	90 - 110	P	03/10/2025	12:27	LB134965
	Zinc	2470	2500	99	90 - 110	P	03/10/2025	12:27	LB134965
CCV02	Aluminum	9920	10000	99	90 - 110	P	03/10/2025	13:21	LB134965
	Antimony	5190	5000	104	90 - 110	P	03/10/2025	13:21	LB134965
	Arsenic	5090	5000	102	90 - 110	P	03/10/2025	13:21	LB134965
	Barium	10200	10000	102	90 - 110	P	03/10/2025	13:21	LB134965
	Beryllium	231	250	93	90 - 110	P	03/10/2025	13:21	LB134965
	Cadmium	2470	2500	99	90 - 110	P	03/10/2025	13:21	LB134965
	Calcium	24300	25000	97	90 - 110	P	03/10/2025	13:21	LB134965
	Chromium	1000	1000	100	90 - 110	P	03/10/2025	13:21	LB134965
	Cobalt	2480	2500	99	90 - 110	P	03/10/2025	13:21	LB134965
	Copper	1270	1250	102	90 - 110	P	03/10/2025	13:21	LB134965
	Iron	5220	5000	104	90 - 110	P	03/10/2025	13:21	LB134965
	Lead	4950	5000	99	90 - 110	P	03/10/2025	13:21	LB134965

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q1422

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1422 **SAS No.:** Q1422

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	24000	25000	96	90 - 110	P	03/10/2025	13:21	LB134965
	Manganese	2420	2500	97	90 - 110	P	03/10/2025	13:21	LB134965
	Nickel	2490	2500	100	90 - 110	P	03/10/2025	13:21	LB134965
	Potassium	26700	25000	107	90 - 110	P	03/10/2025	13:21	LB134965
	Selenium	5150	5000	103	90 - 110	P	03/10/2025	13:21	LB134965
	Silver	1260	1250	101	90 - 110	P	03/10/2025	13:21	LB134965
	Sodium	26400	25000	106	90 - 110	P	03/10/2025	13:21	LB134965
	Thallium	5050	5000	101	90 - 110	P	03/10/2025	13:21	LB134965
	Vanadium	2470	2500	99	90 - 110	P	03/10/2025	13:21	LB134965
	Zinc	2550	2500	102	90 - 110	P	03/10/2025	13:21	LB134965
CCV03	Aluminum	10000	10000	100	90 - 110	P	03/10/2025	14:11	LB134965
	Antimony	5250	5000	105	90 - 110	P	03/10/2025	14:11	LB134965
	Arsenic	5180	5000	104	90 - 110	P	03/10/2025	14:11	LB134965
	Barium	10200	10000	102	90 - 110	P	03/10/2025	14:11	LB134965
	Beryllium	239	250	96	90 - 110	P	03/10/2025	14:11	LB134965
	Cadmium	2500	2500	100	90 - 110	P	03/10/2025	14:11	LB134965
	Calcium	24500	25000	98	90 - 110	P	03/10/2025	14:11	LB134965
	Chromium	1000	1000	100	90 - 110	P	03/10/2025	14:11	LB134965
	Cobalt	2510	2500	100	90 - 110	P	03/10/2025	14:11	LB134965
	Copper	1300	1250	104	90 - 110	P	03/10/2025	14:11	LB134965
	Iron	5200	5000	104	90 - 110	P	03/10/2025	14:11	LB134965
	Lead	5010	5000	100	90 - 110	P	03/10/2025	14:11	LB134965
	Magnesium	24300	25000	97	90 - 110	P	03/10/2025	14:11	LB134965
	Manganese	2420	2500	97	90 - 110	P	03/10/2025	14:11	LB134965
	Nickel	2520	2500	101	90 - 110	P	03/10/2025	14:11	LB134965
	Potassium	26600	25000	106	90 - 110	P	03/10/2025	14:11	LB134965
	Selenium	5240	5000	105	90 - 110	P	03/10/2025	14:11	LB134965
	Silver	1280	1250	102	90 - 110	P	03/10/2025	14:11	LB134965
	Sodium	25900	25000	104	90 - 110	P	03/10/2025	14:11	LB134965
	Thallium	5010	5000	100	90 - 110	P	03/10/2025	14:11	LB134965
	Vanadium	2500	2500	100	90 - 110	P	03/10/2025	14:11	LB134965
	Zinc	2590	2500	104	90 - 110	P	03/10/2025	14:11	LB134965
CCV04	Aluminum	9730	10000	97	90 - 110	P	03/10/2025	15:01	LB134965
	Antimony	5120	5000	102	90 - 110	P	03/10/2025	15:01	LB134965

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q1422

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1422 **SAS No.:** Q1422

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	5010	5000	100	90 - 110	P	03/10/2025	15:01	LB134965
	Barium	9950	10000	100	90 - 110	P	03/10/2025	15:01	LB134965
	Beryllium	226	250	90	90 - 110	P	03/10/2025	15:01	LB134965
	Cadmium	2420	2500	97	90 - 110	P	03/10/2025	15:01	LB134965
	Calcium	24000	25000	96	90 - 110	P	03/10/2025	15:01	LB134965
	Chromium	993	1000	99	90 - 110	P	03/10/2025	15:01	LB134965
	Cobalt	2440	2500	98	90 - 110	P	03/10/2025	15:01	LB134965
	Copper	1250	1250	100	90 - 110	P	03/10/2025	15:01	LB134965
	Iron	5240	5000	105	90 - 110	P	03/10/2025	15:01	LB134965
	Lead	4860	5000	97	90 - 110	P	03/10/2025	15:01	LB134965
	Magnesium	23700	25000	95	90 - 110	P	03/10/2025	15:01	LB134965
	Manganese	2360	2500	94	90 - 110	P	03/10/2025	15:01	LB134965
	Nickel	2440	2500	98	90 - 110	P	03/10/2025	15:01	LB134965
	Potassium	26800	25000	107	90 - 110	P	03/10/2025	15:01	LB134965
	Selenium	5050	5000	101	90 - 110	P	03/10/2025	15:01	LB134965
	Silver	1250	1250	100	90 - 110	P	03/10/2025	15:01	LB134965
	Sodium	26100	25000	104	90 - 110	P	03/10/2025	15:01	LB134965
	Thallium	4780	5000	96	90 - 110	P	03/10/2025	15:01	LB134965
	Vanadium	2450	2500	98	90 - 110	P	03/10/2025	15:01	LB134965
	Zinc	2510	2500	100	90 - 110	P	03/10/2025	15:01	LB134965
CCV05	Aluminum	9800	10000	98	90 - 110	P	03/10/2025	15:51	LB134965
	Antimony	5090	5000	102	90 - 110	P	03/10/2025	15:51	LB134965
	Arsenic	5000	5000	100	90 - 110	P	03/10/2025	15:51	LB134965
	Barium	9730	10000	97	90 - 110	P	03/10/2025	15:51	LB134965
	Beryllium	239	250	96	90 - 110	P	03/10/2025	15:51	LB134965
	Cadmium	2420	2500	97	90 - 110	P	03/10/2025	15:51	LB134965
	Calcium	23800	25000	95	90 - 110	P	03/10/2025	15:51	LB134965
	Chromium	983	1000	98	90 - 110	P	03/10/2025	15:51	LB134965
	Cobalt	2440	2500	98	90 - 110	P	03/10/2025	15:51	LB134965
	Copper	1250	1250	100	90 - 110	P	03/10/2025	15:51	LB134965
	Iron	4890	5000	98	90 - 110	P	03/10/2025	15:51	LB134965
	Lead	4850	5000	97	90 - 110	P	03/10/2025	15:51	LB134965
	Magnesium	24000	25000	96	90 - 110	P	03/10/2025	15:51	LB134965
	Manganese	2360	2500	94	90 - 110	P	03/10/2025	15:51	LB134965

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q1422

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1422 **SAS No.:** Q1422

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2450	2500	98	90 - 110	P	03/10/2025	15:51	LB134965
	Potassium	25000	25000	100	90 - 110	P	03/10/2025	15:51	LB134965
	Selenium	5020	5000	100	90 - 110	P	03/10/2025	15:51	LB134965
	Silver	1250	1250	100	90 - 110	P	03/10/2025	15:51	LB134965
	Sodium	24100	25000	96	90 - 110	P	03/10/2025	15:51	LB134965
	Thallium	4850	5000	97	90 - 110	P	03/10/2025	15:51	LB134965
	Vanadium	2450	2500	98	90 - 110	P	03/10/2025	15:51	LB134965
	Zinc	2500	2500	100	90 - 110	P	03/10/2025	15:51	LB134965
CCV06	Aluminum	9600	10000	96	90 - 110	P	03/10/2025	16:48	LB134965
	Antimony	5090	5000	102	90 - 110	P	03/10/2025	16:48	LB134965
	Arsenic	4980	5000	100	90 - 110	P	03/10/2025	16:48	LB134965
	Barium	9720	10000	97	90 - 110	P	03/10/2025	16:48	LB134965
	Beryllium	268	250	107	90 - 110	P	03/10/2025	16:48	LB134965
	Cadmium	2410	2500	96	90 - 110	P	03/10/2025	16:48	LB134965
	Calcium	23700	25000	95	90 - 110	P	03/10/2025	16:48	LB134965
	Chromium	977	1000	98	90 - 110	P	03/10/2025	16:48	LB134965
	Cobalt	2430	2500	97	90 - 110	P	03/10/2025	16:48	LB134965
	Copper	1250	1250	100	90 - 110	P	03/10/2025	16:48	LB134965
	Iron	5100	5000	102	90 - 110	P	03/10/2025	16:48	LB134965
	Lead	4830	5000	96	90 - 110	P	03/10/2025	16:48	LB134965
	Magnesium	23600	25000	94	90 - 110	P	03/10/2025	16:48	LB134965
	Manganese	2680	2500	107	90 - 110	P	03/10/2025	16:48	LB134965
	Nickel	2440	2500	97	90 - 110	P	03/10/2025	16:48	LB134965
	Potassium	26100	25000	104	90 - 110	P	03/10/2025	16:48	LB134965
	Selenium	5020	5000	100	90 - 110	P	03/10/2025	16:48	LB134965
	Silver	1240	1250	100	90 - 110	P	03/10/2025	16:48	LB134965
	Sodium	24900	25000	100	90 - 110	P	03/10/2025	16:48	LB134965
	Thallium	4890	5000	98	90 - 110	P	03/10/2025	16:48	LB134965
	Vanadium	2420	2500	97	90 - 110	P	03/10/2025	16:48	LB134965
	Zinc	2500	2500	100	90 - 110	P	03/10/2025	16:48	LB134965
CCV07	Aluminum	9380	10000	94	90 - 110	P	03/10/2025	17:39	LB134965
	Antimony	5010	5000	100	90 - 110	P	03/10/2025	17:39	LB134965
	Arsenic	4890	5000	98	90 - 110	P	03/10/2025	17:39	LB134965
	Barium	9380	10000	94	90 - 110	P	03/10/2025	17:39	LB134965

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q1422

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1422 **SAS No.:** Q1422

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	259	250	104	90 - 110	P	03/10/2025	17:39	LB134965
	Cadmium	2350	2500	94	90 - 110	P	03/10/2025	17:39	LB134965
	Calcium	22900	25000	92	90 - 110	P	03/10/2025	17:39	LB134965
	Chromium	942	1000	94	90 - 110	P	03/10/2025	17:39	LB134965
	Cobalt	2370	2500	95	90 - 110	P	03/10/2025	17:39	LB134965
	Copper	1220	1250	98	90 - 110	P	03/10/2025	17:39	LB134965
	Iron	4820	5000	96	90 - 110	P	03/10/2025	17:39	LB134965
	Lead	4710	5000	94	90 - 110	P	03/10/2025	17:39	LB134965
	Magnesium	23000	25000	92	90 - 110	P	03/10/2025	17:39	LB134965
	Manganese	2580	2500	103	90 - 110	P	03/10/2025	17:39	LB134965
	Nickel	2380	2500	95	90 - 110	P	03/10/2025	17:39	LB134965
	Potassium	24600	25000	98	90 - 110	P	03/10/2025	17:39	LB134965
	Selenium	4950	5000	99	90 - 110	P	03/10/2025	17:39	LB134965
	Silver	1190	1250	95	90 - 110	P	03/10/2025	17:39	LB134965
	Sodium	23300	25000	93	90 - 110	P	03/10/2025	17:39	LB134965
	Thallium	4820	5000	96	90 - 110	P	03/10/2025	17:39	LB134965
	Vanadium	2350	2500	94	90 - 110	P	03/10/2025	17:39	LB134965
	Zinc	2410	2500	96	90 - 110	P	03/10/2025	17:39	LB134965
CCV08	Aluminum	9210	10000	92	90 - 110	P	03/10/2025	18:29	LB134965
	Antimony	4880	5000	98	90 - 110	P	03/10/2025	18:29	LB134965
	Arsenic	4780	5000	96	90 - 110	P	03/10/2025	18:29	LB134965
	Barium	9560	10000	96	90 - 110	P	03/10/2025	18:29	LB134965
	Beryllium	259	250	103	90 - 110	P	03/10/2025	18:29	LB134965
	Cadmium	2320	2500	93	90 - 110	P	03/10/2025	18:29	LB134965
	Calcium	22900	25000	92	90 - 110	P	03/10/2025	18:29	LB134965
	Chromium	933	1000	93	90 - 110	P	03/10/2025	18:29	LB134965
	Cobalt	2340	2500	94	90 - 110	P	03/10/2025	18:29	LB134965
	Copper	1200	1250	96	90 - 110	P	03/10/2025	18:29	LB134965
	Iron	4930	5000	99	90 - 110	P	03/10/2025	18:29	LB134965
	Lead	4630	5000	93	90 - 110	P	03/10/2025	18:29	LB134965
	Magnesium	22900	25000	92	90 - 110	P	03/10/2025	18:29	LB134965
	Manganese	2610	2500	104	90 - 110	P	03/10/2025	18:29	LB134965
	Nickel	2350	2500	94	90 - 110	P	03/10/2025	18:29	LB134965
	Potassium	25100	25000	100	90 - 110	P	03/10/2025	18:29	LB134965

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q1422

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1422 **SAS No.:** Q1422

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Selenium	4800	5000	96	90 - 110	P	03/10/2025	18:29	LB134965
	Silver	1200	1250	96	90 - 110	P	03/10/2025	18:29	LB134965
	Sodium	23800	25000	95	90 - 110	P	03/10/2025	18:29	LB134965
	Thallium	4760	5000	95	90 - 110	P	03/10/2025	18:29	LB134965
	Vanadium	2330	2500	93	90 - 110	P	03/10/2025	18:29	LB134965
	Zinc	2410	2500	96	90 - 110	P	03/10/2025	18:29	LB134965
CCV09	Aluminum	9210	10000	92	90 - 110	P	03/10/2025	18:46	LB134965
	Antimony	4910	5000	98	90 - 110	P	03/10/2025	18:46	LB134965
	Arsenic	4800	5000	96	90 - 110	P	03/10/2025	18:46	LB134965
	Barium	9700	10000	97	90 - 110	P	03/10/2025	18:46	LB134965
	Beryllium	253	250	101	90 - 110	P	03/10/2025	18:46	LB134965
	Cadmium	2320	2500	93	90 - 110	P	03/10/2025	18:46	LB134965
	Calcium	22900	25000	92	90 - 110	P	03/10/2025	18:46	LB134965
	Chromium	923	1000	92	90 - 110	P	03/10/2025	18:46	LB134965
	Cobalt	2340	2500	94	90 - 110	P	03/10/2025	18:46	LB134965
	Copper	1210	1250	97	90 - 110	P	03/10/2025	18:46	LB134965
	Iron	4910	5000	98	90 - 110	P	03/10/2025	18:46	LB134965
	Lead	4630	5000	93	90 - 110	P	03/10/2025	18:46	LB134965
	Magnesium	22700	25000	91	90 - 110	P	03/10/2025	18:46	LB134965
	Manganese	2540	2500	102	90 - 110	P	03/10/2025	18:46	LB134965
	Nickel	2350	2500	94	90 - 110	P	03/10/2025	18:46	LB134965
	Potassium	25300	25000	101	90 - 110	P	03/10/2025	18:46	LB134965
	Selenium	4840	5000	97	90 - 110	P	03/10/2025	18:46	LB134965
	Silver	1180	1250	95	90 - 110	P	03/10/2025	18:46	LB134965
	Sodium	24000	25000	96	90 - 110	P	03/10/2025	18:46	LB134965
	Thallium	4750	5000	95	90 - 110	P	03/10/2025	18:46	LB134965
	Vanadium	2340	2500	94	90 - 110	P	03/10/2025	18:46	LB134965
	Zinc	2360	2500	94	90 - 110	P	03/10/2025	18:46	LB134965