

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB166872BS

Contract: PORT06

Lab Code: CHEM Case No.: Q1422 SAS No.: Q1422 SDG NO.: Q1422

Lab Sample ID: PB166872BS Date(s) Analyzed: 02/26/2025 02/26/2025

Instrument ID (1): ECD_P Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP070069.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.677	5.627	5.727	167	171	2.31
COLUMN 1	2	5.698	5.648	5.748	175		
	3	5.761	5.711	5.811	168		
	4	5.858	5.808	5.908	170		
	5	6.151	6.101	6.201	173		
COLUMN 2	1	4.919	4.869	4.969	175	175	
	2	4.938	4.888	4.988	173		
	3	5.115	5.065	5.165	180		
	4	5.157	5.107	5.207	175		
	5	5.372	5.322	5.422	172		
Aroclor-1260	1	7.271	7.221	7.321	186	166	0
COLUMN 1	2	7.524	7.474	7.574	171		
	3	7.884	7.834	7.934	153		
	4	8.108	8.058	8.158	160		
	5	8.428	8.378	8.478	158		
COLUMN 2	1	6.41	6.36	6.46	172	166	
	2	6.598	6.548	6.648	172		
	3	6.752	6.702	6.802	164		
	4	7.225	7.175	7.275	158		
	5	7.466	7.416	7.516	163		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

NB-07-022525MS

Contract: PORT06

Lab Code: CHEM Case No.: Q1422 SAS No.: Q1422 SDG NO.: Q1422

Lab Sample ID: Q1428-01MS Date(s) Analyzed: 02/26/2025 02/26/2025

Instrument ID (1): ECD_P Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP070074.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.676	5.626	5.726	158	152	5.13
COLUMN 1	2	5.698	5.648	5.748	147		
	3	5.76	5.71	5.81	144		
	4	5.858	5.808	5.908	155		
	5	6.15	6.1	6.2	157		
	1	4.919	4.869	4.969	152		
COLUMN 2	2	4.939	4.889	4.989	151	160	
	3	5.116	5.066	5.166	142		
	4	5.158	5.108	5.208	158		
	5	5.373	5.323	5.423	196		
Aroclor-1260	1	7.271	7.221	7.321	166	162	9.71
COLUMN 1	2	7.525	7.475	7.575	198		
	3	7.884	7.834	7.934	137		
	4	8.107	8.057	8.157	159		
	5	8.429	8.379	8.479	149		
	1	6.411	6.361	6.461	162	147	
COLUMN 2	2	6.599	6.549	6.649	159		
	3	6.752	6.702	6.802	151		
	4	7.226	7.176	7.276	132		
	5	7.466	7.416	7.516	132		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

NB-07-022525MSD

Contract: PORT06

Lab Code: CHEM Case No.: Q1422 SAS No.: Q1422 SDG NO.: Q1422

Lab Sample ID: Q1428-01MSD Date(s) Analyzed: 02/26/2025 02/26/2025

Instrument ID (1): ECD_P Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP070075.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.679	5.629	5.729	139	142	10.67
COLUMN 1	2	5.701	5.651	5.751	143		
	3	5.764	5.714	5.814	137		
	4	5.861	5.811	5.911	147		
	5	6.153	6.103	6.203	142		
	1	4.918	4.868	4.968	147		
COLUMN 2	2	4.938	4.888	4.988	151	158	
	3	5.115	5.065	5.165	142		
	4	5.157	5.107	5.207	156		
	5	5.372	5.322	5.422	193		
Aroclor-1260	1	7.274	7.224	7.324	159	156	10.1
COLUMN 1	2	7.528	7.478	7.578	194		
	3	7.886	7.836	7.936	133		
	4	8.11	8.06	8.16	148		
	5	8.431	8.381	8.481	145		
	1	6.41	6.36	6.46	157		
COLUMN 2	2	6.597	6.547	6.647	156	141	
	3	6.751	6.701	6.801	140		
	4	7.224	7.174	7.274	128		
	5	7.465	7.415	7.515	126		