

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1422 SAS No.: Q1422 SDG NO.: Q1422

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL01 Date Analyzed: 02/26/2025

Lab Sample No.: AR1660CCC500 Data File : PP070063.D Time Analyzed: 09:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.682	5.581	5.781	508.150	500.000	1.6
Aroclor-1016-2	5.704	5.602	5.802	531.110	500.000	6.2
Aroclor-1016-3	5.766	5.665	5.865	514.780	500.000	3.0
Aroclor-1016-4	5.863	5.762	5.962	532.360	500.000	6.5
Aroclor-1016-5	6.157	6.056	6.256	527.970	500.000	5.6
Aroclor-1260-1	7.276	7.175	7.375	525.940	500.000	5.2
Aroclor-1260-2	7.530	7.429	7.629	505.490	500.000	1.1
Aroclor-1260-3	7.887	7.788	7.988	575.480	500.000	15.1
Aroclor-1260-4	8.113	8.012	8.212	518.530	500.000	3.7
Aroclor-1260-5	8.434	8.333	8.533	556.170	500.000	11.2
Decachlorobiphenyl	10.259	10.156	10.356	52.830	50.000	5.7
Tetrachloro-m-xylene	4.529	4.427	4.627	55.320	50.000	10.6

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1422 SAS No.: Q1422 SDG NO.: Q1422

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL01 Date Analyzed: 02/26/2025

Lab Sample No.: AR1660CCC500 Data File : PP070063.D Time Analyzed: 09:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.920	4.820	5.020	522.760	500.000	4.6
Aroclor-1016-2	4.939	4.839	5.039	521.230	500.000	4.2
Aroclor-1016-3	5.117	5.017	5.217	539.820	500.000	8.0
Aroclor-1016-4	5.158	5.058	5.258	529.840	500.000	6.0
Aroclor-1016-5	5.373	5.273	5.473	575.230	500.000	15.0
Aroclor-1260-1	6.411	6.312	6.512	534.150	500.000	6.8
Aroclor-1260-2	6.599	6.500	6.700	536.120	500.000	7.2
Aroclor-1260-3	6.754	6.654	6.854	505.810	500.000	1.2
Aroclor-1260-4	7.226	7.126	7.326	554.670	500.000	10.9
Aroclor-1260-5	7.466	7.367	7.567	554.650	500.000	10.9
Decachlorobiphenyl	8.888	8.788	8.988	52.110	50.000	4.2
Tetrachloro-m-xylene	3.831	3.730	3.930	51.760	50.000	3.5