

Surrogate Summary

SDG No.: **Q1422**

Client: **Portal Partners Tri-Venture**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1422-01	B-154-SB01	1,2-Dichloroethane-d4	50	57.4	115	70 (50)	130 (163)
		Dibromofluoromethane	50	55.9	112	70 (54)	130 (147)
		Toluene-d8	50	46.9	94	70 (58)	130 (134)
		4-Bromofluorobenzene	50	25.9	52 *	70 (29)	130 (146)
Q1422-01RE	B-154-SB01RE	1,2-Dichloroethane-d4	50	48.5	97	70 (63)	130 (155)
		Dibromofluoromethane	50	51.9	104	70 (70)	130 (134)
		Toluene-d8	50	46.3	93	70 (74)	130 (123)
		4-Bromofluorobenzene	50	36.8	74	70 (38)	130 (136)
Q1422-02	B-154-SB02	1,2-Dichloroethane-d4	50	56.0	112	70 (63)	130 (155)
		Dibromofluoromethane	50	53.9	108	70 (70)	130 (134)
		Toluene-d8	50	47.5	95	70 (74)	130 (123)
		4-Bromofluorobenzene	50	41.2	82	70 (38)	130 (136)
VY0227SBL01	VY0227SBL01	1,2-Dichloroethane-d4	50	59.2	118	70 (50)	130 (163)
		Dibromofluoromethane	50	52.5	105	70 (54)	130 (147)
		Toluene-d8	50	48.2	96	70 (58)	130 (134)
		4-Bromofluorobenzene	50	49.5	99	70 (29)	130 (146)
VY0227SBS01	VY0227SBS01	1,2-Dichloroethane-d4	50	51.1	102	70 (50)	130 (163)
		Dibromofluoromethane	50	52.3	105	70 (54)	130 (147)
		Toluene-d8	50	52.3	105	70 (58)	130 (134)
		4-Bromofluorobenzene	50	52.4	105	70 (29)	130 (146)
VY0227SBSD01	VY0227SBSD01	1,2-Dichloroethane-d4	50	45.2	90	70 (50)	130 (163)
		Dibromofluoromethane	50	45.6	91	70 (54)	130 (147)
		Toluene-d8	50	45.3	91	70 (58)	130 (134)
		4-Bromofluorobenzene	50	45.3	91	70 (29)	130 (146)
VY0228SBL01	VY0228SBL01	1,2-Dichloroethane-d4	50	52.2	104	70 (63)	130 (155)
		Dibromofluoromethane	50	52.0	104	70 (70)	130 (134)
		Toluene-d8	50	47.1	94	70 (74)	130 (123)
		4-Bromofluorobenzene	50	40.7	81	70 (38)	130 (136)
VY0228SBS01	VY0228SBS01	1,2-Dichloroethane-d4	50	43.2	86	70 (63)	130 (155)
		Dibromofluoromethane	50	45.7	91	70 (70)	130 (134)
		Toluene-d8	50	44.4	89	70 (74)	130 (123)
		4-Bromofluorobenzene	50	43.7	87	70 (38)	130 (136)

