

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q1422 SAS No.: Q1422 SDG NO.: Q1422
 Lab File ID: VY021375.D Date Analyzed: 02/28/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:44
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	170466	7.71	255493	8.62	227302	11.42
UPPER LIMIT	340932	8.207	510986	9.116	454604	11.92
LOWER LIMIT	85233	7.207	127747	8.116	113651	10.92
EPA SAMPLE NO.						
B-154-SB01RE	101641	7.71	171514	8.62	143270	11.41
B-154-SB02	137587	7.71	246433	8.62	216935	11.42
VY0228SBL01	105534	7.71	180558	8.62	159029	11.42
VY0228SBS01	168970	7.71	257471	8.62	224033	11.42

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.