

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: PORT06
Lab Code: CHEM Case No.: Q1422 SAS No.: Q1422 SDG NO.: Q1422
Lab File ID: VY021308.D BFB Injection Date: 02/25/2025
Instrument ID: MSVOA_Y BFB Injection Time: 12:10
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20.8
75	30.0 - 60.0% of mass 95	55.4
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.7
173	Less than 2.0% of mass 174	0.7 (0.8) 1
174	50.0 - 100.0% of mass 95	84.7
175	5.0 - 9.0% of mass 174	6.4 (7.5) 1
176	95.0 - 101.0% of mass 174	83.6 (98.8) 1
177	5.0 - 9.0% of mass 176	5.5 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VY021309.D	02/25/2025	12:40
VSTDIC010	VSTDIC010	VY021310.D	02/25/2025	13:03
VSTDIC020	VSTDIC020	VY021311.D	02/25/2025	13:26
VSTDIC050	VSTDIC050	VY021312.D	02/25/2025	14:48
VSTDIC150	VSTDIC150	VY021314.D	02/25/2025	15:57
VSTDIC100	VSTDIC100	VY021316.D	02/25/2025	16:49