

Surrogate Summary

SW-846

SDG No.: Q1423

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB166876BL	PB166876BL	2-Methylnaphthalene-d10	0.4	0.34	84		30	150
		Fluoranthene-d10	0.4	0.38	94		30	150
		Nitrobenzene-d5	0.4	0.34	86		55	111
		2-Fluorobiphenyl	0.4	0.30	75		53	106
		Terphenyl-d14	0.4	0.45	113		58	132
PB166876BS	PB166876BS	2-Methylnaphthalene-d10	0.4	0.43	108		30	150
		Fluoranthene-d10	0.4	0.34	85		30	150
		Nitrobenzene-d5	0.4	0.38	96		55	111
		2-Fluorobiphenyl	0.4	0.49	123	*	53	106
		Terphenyl-d14	0.4	0.48	119		58	132
PB166876BSD	PB166876BSD	2-Methylnaphthalene-d10	0.4	0.42	106		30	150
		Fluoranthene-d10	0.4	0.34	84		30	150
		Nitrobenzene-d5	0.4	0.39	97		55	111
		2-Fluorobiphenyl	0.4	0.51	127	*	53	106
		Terphenyl-d14	0.4	0.47	118		58	132
Q1423-02	VPB192-HYD-20250221	2-Methylnaphthalene-d10	0.4	0.086	22	*	30	150
		Fluoranthene-d10	0.4	0.39	97		30	150
		Nitrobenzene-d5	0.4	0.33	81		55	111
		2-Fluorobiphenyl	0.4	0.35	87		53	106
		Terphenyl-d14	0.4	0.59	147	*	58	132
Q1423-03	BP-VPB-192-EB-20250221	2-Methylnaphthalene-d10	0.4	0.37	93		30	150
		Fluoranthene-d10	0.4	0.42	105		30	150
		Nitrobenzene-d5	0.4	0.37	91		55	111
		2-Fluorobiphenyl	0.4	0.50	124	*	53	106
		Terphenyl-d14	0.4	0.63	158	*	58	132