

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1423

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified DataFile: BN036513.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB166876BSD	1,4-Dioxane	0.4	0.37	ug/L	93	0			70	130	20