

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1423 SAS No.: Q1423 SDG NO.: Q1423

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 02/26/2025

Lab File ID: BN036508.D Time Analyzed: 13:43

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2772	7.739	6909	10.53	4680	14.38
UPPER LIMIT	5544	8.239	13818	11.03	9360	14.877
LOWER LIMIT	1386	7.239	3454.5	10.03	2340	13.877
EPA SAMPLE NO.						
01 PB166876BL	2585	7.75	5224	10.55	3049	14.39
02 VPB192-HYD-20250221	2028	7.74	4364	10.53	3005	14.38
03 PB166876BS	2531	7.74	5759	10.53	3208	14.38
04 BP-VPB-192-EB-20250221	2561	7.74	5724	10.53	3696	14.38
05 PB166876BSD	2476	7.74	5554	10.53	2910	14.38

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.