

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1423 SAS No.: Q1423 SDG No.: Q1423

Instrument ID: MSVOA_X Calibration Date/Time: 02/25/2025 10:00

Lab File ID: VX045029.D Init. Calib. Date(s): 02/10/2025 02/10/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:25 12:28

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.854	0.745	0.1	-12.76	20
Vinyl Chloride	0.827	0.732		-11.49	20
Bromomethane	0.247	0.252		2.02	20
Chloroethane	0.307	0.346		12.7	20
Trichlorofluoromethane	1.046	1.033		-1.24	20
1,1,2-Trichlorotrifluoroethane	0.632	0.617		-2.37	20
1,1-Dichloroethene	0.644	0.593		-7.92	20
Acetone	0.294	0.364		23.81	20
Carbon Disulfide	1.767	1.454		-17.71	20
Methyl tert-butyl Ether	2.050	1.972		-3.81	20
Methylene Chloride	0.721	0.673		-6.66	20
trans-1,2-Dichloroethene	0.634	0.582		-8.2	20
1,1-Dichloroethane	1.233	1.194	0.1	-3.16	20
2-Butanone	0.478	0.544		13.81	20
Carbon Tetrachloride	0.460	0.478		3.91	20
cis-1,2-Dichloroethene	0.762	0.721		-5.38	20
Chloroform	1.196	1.170		-2.17	20
1,1,1-Trichloroethane	1.014	0.979		-3.45	20
Methylcyclohexane	0.606	0.609		0.5	20
Benzene	1.465	1.459		-0.41	20
1,2-Dichloroethane	0.467	0.504		7.92	20
Trichloroethene	0.335	0.339		1.19	20
1,2-Dichloropropane	0.364	0.376		3.3	20
Bromodichloromethane	0.489	0.538		10.02	20
4-Methyl-2-Pentanone	0.512	0.618		20.7	20
Toluene	0.872	0.882		1.15	20
t-1,3-Dichloropropene	0.495	0.515		4.04	20
cis-1,3-Dichloropropene	0.552	0.576		4.35	20
1,1,2-Trichloroethane	0.335	0.356		6.27	20
2-Hexanone	0.369	0.458		24.12	20
Dibromochloromethane	0.359	0.394		9.75	20
Tetrachloroethene	0.315	0.319		1.27	20
Chlorobenzene	1.074	1.097	0.3	2.14	20
Ethyl Benzene	1.895	1.927		1.69	20
m/p-Xylenes	0.699	0.720		3	20
o-Xylene	0.701	0.716		2.14	20
Styrene	1.130	1.200		6.2	20
Bromoform	0.258	0.297	0.1	15.12	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1423 SAS No.: Q1423 SDG No.: Q1423

Instrument ID: MSVOA_X Calibration Date/Time: 02/25/2025 10:00

Lab File ID: VX045029.D Init. Calib. Date(s): 02/10/2025 02/10/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:25 12:28

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	4.026	3.989		-0.92	20
1,1,2,2-Tetrachloroethane	1.383	1.395	0.3	0.87	20
1,3-Dichlorobenzene	1.678	1.702		1.43	20
1,4-Dichlorobenzene	1.697	1.703		0.35	20
1,2-Dichlorobenzene	1.650	1.713		3.82	20
1,2-Dichloroethane-d4	0.732	0.737		0.68	20
Dibromofluoromethane	0.325	0.348		7.08	20
Toluene-d8	1.229	1.235		0.49	20
4-Bromofluorobenzene	0.414	0.445		7.49	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.