

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1423 SAS No.: Q1423 SDG No.: Q1423

Instrument ID: MSVOA\_X Calibration Date/Time: 02/28/2025 18:50

Lab File ID: VX045098.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

| COMPOUND                       | RRF   | RRF050 | MIN<br>RRF | %D     | MAX%D |
|--------------------------------|-------|--------|------------|--------|-------|
| Chloromethane                  | 0.770 | 0.696  | 0.1        | -9.61  | 50    |
| Vinyl Chloride                 | 0.764 | 0.736  |            | -3.66  | 50    |
| Bromomethane                   | 0.300 | 0.261  |            | -13    | 50    |
| Chloroethane                   | 0.352 | 0.305  |            | -13.35 | 50    |
| Trichlorofluoromethane         | 1.012 | 0.992  |            | -1.98  | 50    |
| 1,1,2-Trichlorotrifluoroethane | 0.581 | 0.602  |            | 3.61   | 50    |
| 1,1-Dichloroethene             | 0.614 | 0.600  |            | -2.28  | 50    |
| Acetone                        | 0.369 | 0.330  |            | -10.57 | 50    |
| Carbon Disulfide               | 1.636 | 1.572  |            | -3.91  | 50    |
| Methyl tert-butyl Ether        | 2.061 | 2.046  |            | -0.73  | 50    |
| Methylene Chloride             | 0.731 | 0.691  |            | -5.47  | 50    |
| trans-1,2-Dichloroethene       | 0.607 | 0.624  |            | 2.8    | 50    |
| 1,1-Dichloroethane             | 1.247 | 1.237  | 0.1        | -0.8   | 50    |
| 2-Butanone                     | 0.555 | 0.536  |            | -3.42  | 50    |
| Carbon Tetrachloride           | 0.465 | 0.475  |            | 2.15   | 50    |
| cis-1,2-Dichloroethene         | 0.749 | 0.756  |            | 0.94   | 50    |
| Chloroform                     | 1.239 | 1.219  |            | -1.61  | 50    |
| 1,1,1-Trichloroethane          | 1.000 | 1.013  |            | 1.3    | 50    |
| Methylcyclohexane              | 0.549 | 0.612  |            | 11.48  | 50    |
| Benzene                        | 1.448 | 1.471  |            | 1.59   | 50    |
| 1,2-Dichloroethane             | 0.521 | 0.520  |            | -0.19  | 50    |
| Trichloroethene                | 0.340 | 0.344  |            | 1.18   | 50    |
| 1,2-Dichloropropane            | 0.372 | 0.372  |            | 0      | 50    |
| Bromodichloromethane           | 0.516 | 0.518  |            | 0.39   | 50    |
| 4-Methyl-2-Pentanone           | 0.587 | 0.582  |            | -0.85  | 50    |
| Toluene                        | 0.849 | 0.879  |            | 3.53   | 50    |
| t-1,3-Dichloropropene          | 0.431 | 0.460  |            | 6.73   | 50    |
| cis-1,3-Dichloropropene        | 0.503 | 0.536  |            | 6.56   | 50    |
| 1,1,2-Trichloroethane          | 0.350 | 0.342  |            | -2.29  | 50    |
| 2-Hexanone                     | 0.425 | 0.430  |            | 1.18   | 50    |
| Dibromochloromethane           | 0.366 | 0.364  |            | -0.55  | 50    |
| Tetrachloroethene              | 0.320 | 0.334  |            | 4.38   | 50    |
| Chlorobenzene                  | 1.058 | 1.084  | 0.3        | 2.46   | 50    |
| Ethyl Benzene                  | 1.843 | 1.942  |            | 5.37   | 50    |
| m/p-Xylenes                    | 0.675 | 0.717  |            | 6.22   | 50    |
| o-Xylene                       | 0.681 | 0.694  |            | 1.91   | 50    |
| Styrene                        | 1.101 | 1.166  |            | 5.9    | 50    |
| Bromoform                      | 0.266 | 0.275  | 0.1        | 3.38   | 50    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1423 SAS No.: Q1423 SDG No.: Q1423

Instrument ID: MSVOA\_X Calibration Date/Time: 02/28/2025 18:50

Lab File ID: VX045098.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

| COMPOUND                  | RRF   | RRF050 | MIN<br>RRF | %D    | MAX%D |
|---------------------------|-------|--------|------------|-------|-------|
| Isopropylbenzene          | 3.903 | 4.046  |            | 3.66  | 50    |
| 1,1,2,2-Tetrachloroethane | 1.432 | 1.422  | 0.3        | -0.7  | 50    |
| 1,3-Dichlorobenzene       | 1.643 | 1.677  |            | 2.07  | 50    |
| 1,4-Dichlorobenzene       | 1.662 | 1.661  |            | -0.06 | 50    |
| 1,2-Dichlorobenzene       | 1.648 | 1.681  |            | 2     | 50    |
| 1,2-Dichloroethane-d4     | 0.795 | 0.780  |            | -1.89 | 50    |
| Dibromofluoromethane      | 0.334 | 0.338  |            | 1.2   | 50    |
| Toluene-d8                | 1.212 | 1.211  |            | -0.08 | 50    |
| 4-Bromofluorobenzene      | 0.402 | 0.392  |            | -2.49 | 50    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.