

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1423 SAS No.: Q1423 SDG No.: Q1423

Instrument ID: MSVOA\_X Calibration Date/Time: 02/25/2025 16:28

Lab File ID: VX045042.D Init. Calib. Date(s): 02/10/2025 02/10/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:25 12:28

GC Column: DB-624UI ID: 0.18 (mm)

| COMPOUND     | Amount | Conc. | %Dev    | MAX%D |
|--------------|--------|-------|---------|-------|
| Chloroethane | 50.000 | 65.67 | -31.334 | 20    |