

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX021025

Review By	John Carlone	Review On	2/11/2025 9:43:42 AM
Supervise By	Maresh Dadoda	Supervise On	2/11/2025 2:05:25 PM
SubDirectory	VX021025	HP Acquire Method	HP Processing Method 82X021025W.M

STD. NAME	STD REF.#
Tune/Reschk	VP132955
Initial Calibration Stds	VP132960,VP132961,VP132962,VP132963,VP132964,VP132965
CCC	VP132956
Internal Standard/PEM	
ICV/I.BLK	VP132966
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX044867.D	10 Feb 2025 09:35		JC/MD	Ok
2	VSTDICC001	VSTDICC001	VX044868.D	10 Feb 2025 10:25		JC/MD	Ok,M
3	VSTDICC005	VSTDICC005	VX044869.D	10 Feb 2025 10:48		JC/MD	Ok,M
4	VSTDICC020	VSTDICC020	VX044870.D	10 Feb 2025 11:11		JC/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VX044871.D	10 Feb 2025 11:34		JC/MD	Ok,M
6	VSTDICC100	VSTDICC100	VX044872.D	10 Feb 2025 12:05		JC/MD	Ok,M
7	VSTDICC150	VSTDICC150	VX044873.D	10 Feb 2025 12:28		JC/MD	Ok,M
8	IBLK	IBLK	VX044874.D	10 Feb 2025 12:51		JC/MD	Ok
9	VSTDICV050	ICVVX021025	VX044875.D	10 Feb 2025 13:15		JC/MD	Ok,M
10	VX0210MBL01	VX0210MBL01	VX044876.D	10 Feb 2025 13:43		JC/MD	Ok
11	VX0210WBL01	VX0210WBL01	VX044877.D	10 Feb 2025 14:06		JC/MD	Ok
12	VX0210WBS01	VX0210WBS01	VX044878.D	10 Feb 2025 14:29		JC/MD	Ok,M
13	VX0210WBSD01	VX0210WBSD01	VX044879.D	10 Feb 2025 14:55		JC/MD	Ok,M
14	Q1250-02	BP-VPB-192-GW-420-4	VX044880.D	10 Feb 2025 15:18	vial B pH<2	JC/MD	Ok
15	Q1250-01	BP-VPB-192-TB-20250	VX044881.D	10 Feb 2025 15:41	vial B pH<2 TB	JC/MD	Ok
16	Q1250-03	BP-VPB-192-GW-300-3	VX044882.D	10 Feb 2025 16:04	vial B pH<2	JC/MD	Ok
17	Q1250-04	BP-VPB-192-GW-320-3	VX044883.D	10 Feb 2025 16:27	vial B pH<2	JC/MD	Ok
18	Q1250-05	BP-VPB-192-GW-340-3	VX044884.D	10 Feb 2025 16:51	vial B pH<2	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX021025

Review By	John Carlone	Review On	2/11/2025 9:43:42 AM
Supervise By	Maresh Dadoda	Supervise On	2/11/2025 2:05:25 PM
SubDirectory	VX021025	HP Acquire Method	HP Processing Method 82X021025W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP132955		
Initial Calibration Stds	VP132960,VP132961,VP132962,VP132963,VP132964,VP132965		
CCC	VP132956		
Internal Standard/PEM			
ICV/I.BLK	VP132966		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q1250-10	BP-VPB-192-GW-380-3	VX044885.D	10 Feb 2025 17:14	vial B pH<2	JC/MD	Ok
20	Q1250-11	BP-VPB-192-GW-400-4	VX044886.D	10 Feb 2025 17:37	vial B pH<2	JC/MD	Ok
21	Q1250-12	BP-VPB-192-GW-440-4	VX044887.D	10 Feb 2025 18:00	vial B pH<2	JC/MD	Ok
22	Q1250-06	BP-VPB-192-GW-360-3	VX044888.D	10 Feb 2025 18:23	vial B pH<2	JC/MD	Ok
23	Q1250-07MS	BP-VPB-192-GW-360-3	VX044889.D	10 Feb 2025 18:46	vial B pH<2	JC/MD	Ok,M
24	Q1250-08MSD	BP-VPB-192-GW-360-3	VX044890.D	10 Feb 2025 19:09	vial B pH<2	JC/MD	Ok,M
25	VSTDCCC050	VSTDCCC050EC	VX044891.D	10 Feb 2025 19:32		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX022525

Review By	John Carlone	Review On	2/26/2025 9:50:41 AM
Supervise By	Maresh Dadoda	Supervise On	2/26/2025 1:46:06 PM
SubDirectory	VX022525	HP Acquire Method	HP Processing Method 82X021025W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133138
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133139,VP133141

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045028.D	25 Feb 2025 09:30		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX045029.D	25 Feb 2025 10:00	CCAL failed high for comp. #16	JC/MD	Ok,M
3	VX0225MBL01	VX0225MBL01	VX045030.D	25 Feb 2025 11:44		JC/MD	Ok
4	VX0225WBL01	VX0225WBL01	VX045031.D	25 Feb 2025 12:07		JC/MD	Ok
5	VX0225WBS01	VX0225WBS01	VX045032.D	25 Feb 2025 12:30		JC/MD	Ok,M
6	VX0225WBSD01	VX0225WBSD01	VX045033.D	25 Feb 2025 12:57		JC/MD	Ok,M
7	IBLK	IBLK	VX045034.D	25 Feb 2025 13:21		JC/MD	Ok
8	Q1401-01	BP-VPB-192-TB-20250	VX045035.D	25 Feb 2025 13:44	vial A pH<2 TB	JC/MD	Ok
9	Q1423-01	BP-VPB-192-TB-20250	VX045036.D	25 Feb 2025 14:07	vial A pH<2 CCAL failed high for comp. #16; TB;Hit of comp.#16	JC/MD	ReRun
10	Q1423-03	BP-VPB-192-EB-20250	VX045037.D	25 Feb 2025 14:30	vial A pH<2 CCAL failed high for comp. #16; EB;Hit of comp.#16	JC/MD	ReRun
11	Q1423-02	VPB192-HYD-2025022	VX045038.D	25 Feb 2025 14:54	vial A pH<2	JC/MD	Ok
12	Q1401-03	BP-VPB-192-GW-840-8	VX045039.D	25 Feb 2025 15:17	vial A pH<2 CCAL failed high for comp. #16	JC/MD	ReRun
13	Q1401-06	BP-VPB-192-GW-900-9	VX045040.D	25 Feb 2025 15:40	vial A pH<2 CCAL failed high for comp. #16	JC/MD	ReRun
14	Q1423-04	BP-VPB-192-GW-925-9	VX045041.D	25 Feb 2025 16:03	vial A pH<2	JC/MD	Ok
15	VSTDCCC050	VSTDCCC050EC	VX045042.D	25 Feb 2025 16:28		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX022825

Review By	Mahesh Dadoda	Review On	2/28/2025 11:09:50 AM
Supervise By	Semsettin Yesilyurt	Supervise On	2/28/2025 11:11:09 AM
SubDirectory	VX022825	HP Acquire Method	HP Processing Method 82X022825W.M

STD. NAME	STD REF.#
Tune/Reschk	VP133187,VP133189
Initial Calibration Stds	VP133194,VP133195,VP133196,VP133197,VP133198,VP133199
CCC	VP133188,VP133190
Internal Standard/PEM	
ICV/I.BLK	VP133200
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045067.D	28 Feb 2025 01:03		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX045068.D	28 Feb 2025 01:27		JC/MD	Ok,M
3	VSTDIC005	VSTDIC005	VX045069.D	28 Feb 2025 02:13		JC/MD	Ok,M
4	VSTDIC020	VSTDIC020	VX045070.D	28 Feb 2025 02:37		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX045071.D	28 Feb 2025 03:00		JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX045072.D	28 Feb 2025 03:23		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX045073.D	28 Feb 2025 03:47		JC/MD	Ok,M
8	IBLK	IBLK	VX045074.D	28 Feb 2025 04:10		JC/MD	Ok
9	VSTDICV050	ICVVX022825	VX045075.D	28 Feb 2025 04:33		JC/MD	Ok,M
10	BFB	BFB	VX045076.D	28 Feb 2025 10:03		JC/MD	Ok
11	VSTDIC050	VSTDIC050	VX045077.D	28 Feb 2025 10:32	V12668	JC/MD	Ok,M
12	VX0228MBL01	VX0228MBL01	VX045078.D	28 Feb 2025 11:00		JC/MD	Ok
13	VX0228WBL01	VX0228WBL01	VX045079.D	28 Feb 2025 11:23		JC/MD	Ok
14	VX0228WBS01	VX0228WBS01	VX045080.D	28 Feb 2025 11:46		JC/MD	Ok,M
15	VX0228WBSD01	VX0228WBSD01	VX045081.D	28 Feb 2025 12:13		JC/MD	Ok,M
16	Q1401-03	BP-VPB-192-GW-840-8	VX045082.D	28 Feb 2025 12:37	vial B pH<2	JC/MD	Ok
17	Q1401-06	BP-VPB-192-GW-900-9	VX045083.D	28 Feb 2025 13:00	vial B pH<2	JC/MD	Ok
18	Q1423-01	BP-VPB-192-TB-20250	VX045084.D	28 Feb 2025 13:23	vial B pH<2 TB	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX022825

Review By	Maresh Dadoda	Review On	2/28/2025 11:09:50 AM
Supervise By	Semsettin Yesilyurt	Supervise On	2/28/2025 11:11:09 AM
SubDirectory	VX022825	HP Acquire Method	HP Processing Method 82X022825W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133187,VP133189		
Initial Calibration Stds	VP133194,VP133195,VP133196,VP133197,VP133198,VP133199		
CCC	VP133188,VP133190		
Internal Standard/PEM			
ICV/I.BLK	VP133200		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q1423-03	BP-VPB-192-EB-20250	VX045085.D	28 Feb 2025 13:47	vial B pH<2 EB	JC/MD	Ok
20	Q1403-01	Storage-Blank-SOIL-RE	VX045086.D	28 Feb 2025 14:10	vial B pH<2	JC/MD	Ok
21	Q1403-02	Storage-Blank-WATER-	VX045087.D	28 Feb 2025 14:33	vial B pH<2	JC/MD	Ok
22	Q1435-01	286107	VX045088.D	28 Feb 2025 14:57	vial A pH<2	JC/MD	Ok
23	Q1403-04	Storage-Blank-SAMPLE	VX045089.D	28 Feb 2025 15:20	vial B pH<2	JC/MD	Ok
24	Q1462-02	FB	VX045090.D	28 Feb 2025 15:44	vial A pH<2 FB	JC/MD	Ok
25	Q1469-01	Storage-Blank-SOIL-RE	VX045091.D	28 Feb 2025 16:07	vial A pH<2	JC/MD	Ok
26	Q1469-02	Storage-Blank-WATER-	VX045092.D	28 Feb 2025 16:31	vial A pH<2	JC/MD	Ok
27	Q1469-03	Storage-Blank-WATER-	VX045093.D	28 Feb 2025 16:54	vial A pH<2	JC/MD	Ok
28	Q1469-04	Storage-Blank-SAMPLE	VX045094.D	28 Feb 2025 17:17	vial A pH<2	JC/MD	Ok
29	Q1403-03	Storage-Blank-WATER-	VX045095.D	28 Feb 2025 17:41	vial B pH<2	JC/MD	Ok
30	Q1462-01	MW2	VX045096.D	28 Feb 2025 18:04	vial A pH<2	JC/MD	Ok,M
31	IBLK	IBLK	VX045097.D	28 Feb 2025 18:27		JC/MD	Ok
32	VSTDCCC050	VSTDCCC050EC	VX045098.D	28 Feb 2025 18:50		JC/MD	Ok,M

M : Manual Integration