

Surrogate Summary

SDG No.: **Q1423**

Client: **Tetra Tech NUS, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1423-01	BP-VPB-192-TB-20250220	1,2-Dichloroethane-d4	50	52.2	104	81	118
		Dibromofluoromethane	50	53.1	106	80	119
		Toluene-d8	50	51.9	104	89	112
		4-Bromofluorobenzene	50	52.3	105	85	114
Q1423-02	VPB192-HYD-20250221	1,2-Dichloroethane-d4	50	55.7	111	81	118
		Dibromofluoromethane	50	53.0	106	80	119
		Toluene-d8	50	50.4	101	89	112
		4-Bromofluorobenzene	50	50.7	101	85	114
Q1423-03	BP-VPB-192-EB-20250221	1,2-Dichloroethane-d4	50	51.2	102	81	118
		Dibromofluoromethane	50	50.4	101	80	119
		Toluene-d8	50	51.4	103	89	112
		4-Bromofluorobenzene	50	52.3	104	85	114
Q1423-04	BP-VPB-192-GW-925-927	1,2-Dichloroethane-d4	50	56.4	113	81	118
		Dibromofluoromethane	50	51.7	103	80	119
		Toluene-d8	50	50.1	100	89	112
		4-Bromofluorobenzene	50	48.9	98	85	114
VX0225WBL01	VX0225WBL01	1,2-Dichloroethane-d4	50	53.6	107	81	118
		Dibromofluoromethane	50	50.3	101	80	119
		Toluene-d8	50	48.4	97	89	112
		4-Bromofluorobenzene	50	47.5	95	85	114
VX0225WBS01	VX0225WBS01	1,2-Dichloroethane-d4	50	53.0	106	81	118
		Dibromofluoromethane	50	52.8	106	80	119
		Toluene-d8	50	50.5	101	89	112
		4-Bromofluorobenzene	50	51.6	103	85	114
VX0225WBSD0	VX0225WBSD01	1,2-Dichloroethane-d4	50	53.4	107	81	118
		Dibromofluoromethane	50	53.7	107	80	119
		Toluene-d8	50	49.7	99	89	112
		4-Bromofluorobenzene	50	53.8	108	85	114
VX0228WBL01	VX0228WBL01	1,2-Dichloroethane-d4	50	50.0	100	81	118
		Dibromofluoromethane	50	51.0	102	80	119
		Toluene-d8	50	50.9	102	89	112
		4-Bromofluorobenzene	50	53.3	107	85	114
VX0228WBS01	VX0228WBS01	1,2-Dichloroethane-d4	50	46.9	94	81	118
		Dibromofluoromethane	50	48.2	96	80	119
		Toluene-d8	50	50.1	100	89	112
		4-Bromofluorobenzene	50	51.0	102	85	114