

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1423

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Datafile : VX045032.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0225WBS01	Chloromethane	20	15.3	ug/L	77			50	139	
	Vinyl chloride	20	16.6	ug/L	83			58	137	
	Bromomethane	20	20.6	ug/L	103			53	141	
	Chloroethane	20	21.1	ug/L	106			60	138	
	Trichlorofluoromethane	20	18.6	ug/L	93			65	141	
	1,1,2-Trichlorotrifluoroethane	20	19.1	ug/L	96			70	136	
	1,1-Dichloroethene	20	17.0	ug/L	85			71	131	
	Acetone	100	110	ug/L	110			39	160	
	Carbon disulfide	20	15.3	ug/L	77			64	133	
	Methyl tert-butyl Ether	20	18.8	ug/L	94			71	124	
	Methylene Chloride	20	18.5	ug/L	93			74	124	
	trans-1,2-Dichloroethene	20	17.7	ug/L	89			75	124	
	1,1-Dichloroethane	20	18.7	ug/L	94			77	125	
	2-Butanone	100	120	ug/L	120			56	143	
	Carbon Tetrachloride	20	19.6	ug/L	98			72	136	
	cis-1,2-Dichloroethene	20	18.8	ug/L	94			78	123	
	Chloroform	20	19.6	ug/L	98			79	124	
	1,1,1-Trichloroethane	20	18.7	ug/L	94			74	131	
	Methylcyclohexane	20	18.2	ug/L	91			72	132	
	Benzene	20	19.1	ug/L	96			79	120	
	1,2-Dichloroethane	20	21.3	ug/L	106			73	128	
	Trichloroethene	20	19.2	ug/L	96			79	123	
	1,2-Dichloropropane	20	19.8	ug/L	99			78	122	
	Bromodichloromethane	20	20.4	ug/L	102			79	125	
	4-Methyl-2-Pentanone	100	120	ug/L	120			67	130	
	Toluene	20	19.6	ug/L	98			80	121	
	t-1,3-Dichloropropene	20	18.7	ug/L	94			73	127	
	cis-1,3-Dichloropropene	20	19.3	ug/L	97			75	124	
	1,1,2-Trichloroethane	20	20.7	ug/L	104			80	119	
	2-Hexanone	100	120	ug/L	120			57	139	
	Dibromochloromethane	20	20.2	ug/L	101			74	126	
	Tetrachloroethene	20	19.2	ug/L	96			74	129	
	Chlorobenzene	20	19.5	ug/L	98			82	118	
	Ethyl Benzene	20	19.0	ug/L	95			79	121	
	m/p-Xylenes	40	38.8	ug/L	97			80	121	
	o-Xylene	20	19.5	ug/L	98			78	122	
	Styrene	20	20.0	ug/L	100			78	123	
	Bromoform	20	21.0	ug/L	105			66	130	
	Isopropylbenzene	20	19.0	ug/L	95			72	131	
	1,1,2,2-Tetrachloroethane	20	19.7	ug/L	99			71	121	
	1,3-Dichlorobenzene	20	18.6	ug/L	93			80	119	
	1,4-Dichlorobenzene	20	18.6	ug/L	93			79	118	
	1,2-Dichlorobenzene	20	19.2	ug/L	96			80	119	