

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1423 SAS No.: Q1423 SDG NO.: Q1423
Lab File ID: VX045029.D Date Analyzed: 02/25/2025
Instrument ID: MSVOA_X Time Analyzed: 10:00
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	110252	5.54	188741	6.75	166334	10.05
UPPER LIMIT	220504	6.044	377482	7.251	332668	10.549
LOWER LIMIT	55126	5.044	94370.5	6.251	83167	9.549
EPA SAMPLE NO.						
VPB192-HYD-20250221	78098	5.54	154583	6.76	143564	10.05
BP-VPB-192-GW-925-927	78013	5.54	157287	6.76	140409	10.06
VX0225WBL01	79395	5.55	159366	6.76	142850	10.05
VX0225WBS01	112915	5.55	200483	6.76	175398	10.05
VX0225WBSD01	104015	5.55	184616	6.76	164218	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.