

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: Q1423 SAS No.: Q1423 SDG NO.: Q1423
 Lab File ID: VX045077.D Date Analyzed: 02/28/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:32
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	76314	12.018				
UPPER LIMIT	152628	12.518				
LOWER LIMIT	38157	11.518				
EPA SAMPLE NO.						
BP-VPB-192-TB-20250220	54116	12.02				
BP-VPB-192-EB-20250221	51925	12.02				
VX0228WBL01	59690	12.02				
VX0228WBS01	76472	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.