

Initial and Continuing Calibration Verification

Client: American Wax

SDG No.: Q1424

Project: Semi Annual 2025

RunNo.: LB134856

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV COD	mg/L	49.329	50	99	95-105	01/22/2025
Sample ID: CCV1 COD	mg/L	50.319	50	101	95-105	02/28/2025
Sample ID: CCV2 COD	mg/L	50.319	50	101	95-105	02/28/2025
Sample ID: CCV3 COD	mg/L	50.319	50	101	95-105	02/28/2025