

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tris Pharma, Inc. **SDG No.:** Q1429
Contract: TRIS02 **Lab Code:** CHEM **Case No.:** Q1429 **SAS No.:** Q1429
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	515	510	101	95 - 105	P	03/03/2025	19:17	LB134881
	Zinc	990	1000	99	95 - 105	P	03/03/2025	19:17	LB134881

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tris Pharma, Inc. **SDG No.:** Q1429
Contract: TRIS02 **Lab Code:** CHEM **Case No.:** Q1429 **SAS No.:** Q1429
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	22.0	20.0	110	80 - 120	P	03/03/2025	19:21	LB134881
	Zinc	40.2	40.0	100	80 - 120	P	03/03/2025	19:21	LB134881

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tris Pharma, Inc. **SDG No.:** Q1429

Contract: TRIS02 **Lab Code:** CHEM **Case No.:** Q1429 **SAS No.:** Q1429

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1190	1250	95	90 - 110	P	03/03/2025	19:57	LB134881
	Zinc	2390	2500	96	90 - 110	P	03/03/2025	19:57	LB134881
CCV02	Copper	1330	1250	106	90 - 110	P	03/03/2025	21:12	LB134881
	Zinc	2630	2500	105	90 - 110	P	03/03/2025	21:12	LB134881
CCV03	Copper	1290	1250	103	90 - 110	P	03/03/2025	22:02	LB134881
	Zinc	2550	2500	102	90 - 110	P	03/03/2025	22:02	LB134881
CCV04	Copper	1250	1250	100	90 - 110	P	03/03/2025	23:07	LB134881
	Zinc	2460	2500	98	90 - 110	P	03/03/2025	23:07	LB134881
CCV05	Copper	1230	1250	98	90 - 110	P	03/04/2025	00:03	LB134881
	Zinc	2420	2500	97	90 - 110	P	03/04/2025	00:03	LB134881

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tris Pharma, Inc. **SDG No.:** Q1429
Contract: TRIS02 **Lab Code:** CHEM **Case No.:** Q1429 **SAS No.:** Q1429
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	524	510	103	95 - 105	P	03/10/2025	11:48	LB134965
	Zinc	1040	1000	104	95 - 105	P	03/10/2025	11:48	LB134965

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tris Pharma, Inc. **SDG No.:** Q1429
Contract: TRIS02 **Lab Code:** CHEM **Case No.:** Q1429 **SAS No.:** Q1429
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	21.2	20.0	106	80 - 120	P	03/10/2025	11:52	LB134965
	Zinc	40.5	40.0	101	80 - 120	P	03/10/2025	11:52	LB134965

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tris Pharma, Inc. SDG No.: Q1429
Contract: TRIS02 Lab Code: CHEM Case No.: Q1429 SAS No.: Q1429
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1240	1250	99	90 - 110	P	03/10/2025	12:27	LB134965
	Zinc	2470	2500	99	90 - 110	P	03/10/2025	12:27	LB134965
CCV02	Copper	1270	1250	102	90 - 110	P	03/10/2025	13:21	LB134965
	Zinc	2550	2500	102	90 - 110	P	03/10/2025	13:21	LB134965
CCV03	Copper	1300	1250	104	90 - 110	P	03/10/2025	14:11	LB134965
	Zinc	2590	2500	104	90 - 110	P	03/10/2025	14:11	LB134965
CCV04	Copper	1250	1250	100	90 - 110	P	03/10/2025	15:01	LB134965
	Zinc	2510	2500	100	90 - 110	P	03/10/2025	15:01	LB134965
CCV05	Copper	1250	1250	100	90 - 110	P	03/10/2025	15:51	LB134965
	Zinc	2500	2500	100	90 - 110	P	03/10/2025	15:51	LB134965
CCV06	Copper	1250	1250	100	90 - 110	P	03/10/2025	16:48	LB134965
	Zinc	2500	2500	100	90 - 110	P	03/10/2025	16:48	LB134965
CCV07	Copper	1220	1250	98	90 - 110	P	03/10/2025	17:39	LB134965
	Zinc	2410	2500	96	90 - 110	P	03/10/2025	17:39	LB134965
CCV08	Copper	1200	1250	96	90 - 110	P	03/10/2025	18:29	LB134965
	Zinc	2410	2500	96	90 - 110	P	03/10/2025	18:29	LB134965
CCV09	Copper	1210	1250	97	90 - 110	P	03/10/2025	18:46	LB134965
	Zinc	2360	2500	94	90 - 110	P	03/10/2025	18:46	LB134965