



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tris Pharma, Inc. **SDG No.:** Q1429
Contract: TRIS02 **Lab Code:** CHEM **Case No.:** Q1429 **SAS No.:** Q1429
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Copper	21.8	20.0	109	40 - 160	P	03/03/2025	19:30	LB134881
	Zinc	40.7	40.0	102	40 - 160	P	03/03/2025	19:30	LB134881
CRI01	Copper	21.4	20.0	107	40 - 160	P	03/10/2025	12:01	LB134965
	Zinc	41.3	40.0	103	40 - 160	P	03/10/2025	12:01	LB134965