



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tris Pharma, Inc.

SDG No.: Q1429

Contract: TRIS02

Lab Code: CHEM

Case No.: Q1429

SAS No.: Q1429

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Copper	20.0	+/-20.0	U	20.0	P	03/03/2025	19:25	LB134881
	Zinc	40.0	+/-40.0	U	40.0	P	03/03/2025	19:25	LB134881

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Tris Pharma, Inc.</u>		SDG No.: <u>Q1429</u>							
Contract: <u>TRIS02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1429</u>	SAS No.: <u>Q1429</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Copper	20.0	+/-20.0	U	20.0	P	03/03/2025	20:01	LB134881
	Zinc	8.86	+/-40.0	J	40.0	P	03/03/2025	20:01	LB134881
CCB02	Copper	20.0	+/-20.0	U	20.0	P	03/03/2025	21:16	LB134881
	Zinc	40.0	+/-40.0	U	40.0	P	03/03/2025	21:16	LB134881
CCB03	Copper	20.0	+/-20.0	U	20.0	P	03/03/2025	22:24	LB134881
	Zinc	40.0	+/-40.0	U	40.0	P	03/03/2025	22:24	LB134881
CCB04	Copper	20.0	+/-20.0	U	20.0	P	03/03/2025	23:11	LB134881
	Zinc	40.0	+/-40.0	U	40.0	P	03/03/2025	23:11	LB134881
CCB05	Copper	20.0	+/-20.0	U	20.0	P	03/04/2025	00:08	LB134881
	Zinc	40.0	+/-40.0	U	40.0	P	03/04/2025	00:08	LB134881

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tris Pharma, Inc.

SDG No.: Q1429

Contract: TRIS02

Lab Code: CHEM

Case No.: Q1429

SAS No.: Q1429

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Copper	20.0	+/-20.0	U	20.0	P	03/10/2025	11:57	LB134965
	Zinc	40.0	+/-40.0	U	40.0	P	03/10/2025	11:57	LB134965

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Tris Pharma, Inc.</u>		SDG No.: <u>Q1429</u>							
Contract: <u>TRIS02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1429</u>	SAS No.: <u>Q1429</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Copper	20.0	+/-20.0	U	20.0	P	03/10/2025	12:34	LB134965
	Zinc	40.0	+/-40.0	U	40.0	P	03/10/2025	12:34	LB134965
CCB02	Copper	20.0	+/-20.0	U	20.0	P	03/10/2025	13:25	LB134965
	Zinc	40.0	+/-40.0	U	40.0	P	03/10/2025	13:25	LB134965
CCB03	Copper	20.0	+/-20.0	U	20.0	P	03/10/2025	14:15	LB134965
	Zinc	40.0	+/-40.0	U	40.0	P	03/10/2025	14:15	LB134965
CCB04	Copper	20.0	+/-20.0	U	20.0	P	03/10/2025	15:05	LB134965
	Zinc	40.0	+/-40.0	U	40.0	P	03/10/2025	15:05	LB134965
CCB05	Copper	20.0	+/-20.0	U	20.0	P	03/10/2025	15:55	LB134965
	Zinc	40.0	+/-40.0	U	40.0	P	03/10/2025	15:55	LB134965
CCB06	Copper	20.0	+/-20.0	U	20.0	P	03/10/2025	16:52	LB134965
	Zinc	40.0	+/-40.0	U	40.0	P	03/10/2025	16:52	LB134965
CCB07	Copper	20.0	+/-20.0	U	20.0	P	03/10/2025	17:44	LB134965
	Zinc	40.0	+/-40.0	U	40.0	P	03/10/2025	17:44	LB134965
CCB08	Copper	20.0	+/-20.0	U	20.0	P	03/10/2025	18:33	LB134965
	Zinc	40.0	+/-40.0	U	40.0	P	03/10/2025	18:33	LB134965
CCB09	Copper	20.0	+/-20.0	U	20.0	P	03/10/2025	18:50	LB134965
	Zinc	40.0	+/-40.0	U	40.0	P	03/10/2025	18:50	LB134965

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tris Pharma, Inc.

SDG No.: Q1429

Contract: TRIS02

Lab Code: CHEM

Case No.: Q1429

SAS No.: Q1429

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------