

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Tris Pharma, Inc.</u>	SDG No.: <u>Q1429</u>
Contract: <u>TRIS02</u>	Lab Code: <u>CHEM</u>
ICS Source: <u>EPA</u>	Case No.: <u>Q1429</u>
	SAS No.: <u>Q1429</u>
	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	-1.27	2.0	64	-18	22	03/03/2025	19:34	LB134881
	Zinc	2.89			-40	40	03/03/2025	19:34	LB134881
ICSAB01	Copper	478	511	94	434	588	03/03/2025	19:38	LB134881
	Zinc	984	952	103	809	1095	03/03/2025	19:38	LB134881
ICSA	Copper	3.20	2.0	160	-18	22	03/03/2025	19:42	LB134881
	Zinc	-2.80			-40	40	03/03/2025	19:42	LB134881
ICSAB	Copper	503	511	98	434	588	03/03/2025	19:46	LB134881
	Zinc	997	952	105	809	1095	03/03/2025	19:46	LB134881
ICSA01	Copper	9.85	2.0	492	-18	22	03/10/2025	12:05	LB134965
	Zinc	2.96			-40	40	03/10/2025	12:05	LB134965
ICSAB01	Copper	518	511	101	434	588	03/10/2025	12:10	LB134965
	Zinc	1070	952	112	809	1095	03/10/2025	12:10	LB134965
ICSA	Copper	3.14	2.0	157	-18	22	03/10/2025	12:14	LB134965
	Zinc	49.6			-40	40	03/10/2025	12:14	LB134965
ICSAB	Copper	531	511	104	434	588	03/10/2025	12:18	LB134965
	Zinc	1060	952	111	809	1095	03/10/2025	12:18	LB134965