

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Tris Pharma, Inc. **level:** low **sdg no.:** Q1429
contract: TRIS02 **lab code:** CHEM **case no.:** Q1429 **sas no.:** Q1429
matrix: Water **sample id:** Q1429-01 **client id:** OUTFALL-DSN-001MS
Percent Solids for Sample: NA **Spiked ID:** Q1429-01MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	234		61.6		150	115		P
Zinc	ug/L	75 - 125	130		12.9	J	100	117		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Tris Pharma, Inc.</u>	level: <u>low</u>	sdg no.: <u>Q1429</u>
contract: <u>TRIS02</u>	lab code: <u>CHEM</u>	case no.: <u>Q1429</u> sas no.: <u>Q1429</u>
matrix: <u>Water</u>	sample id: <u>Q1429-01</u>	client id: <u>OUTFALL-DSN-001MSD</u>
Percent Solids for Sample: NA	Spiked ID: Q1429-01MSD	Percent Solids for Spike Sample: NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	231		61.6		150	113		P
Zinc	ug/L	75 - 125	127		12.9	J	100	114		P