

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TRIS02

Lab Code: CHEM Case No.: Q1429 SAS No.: Q1429 SDG No.: Q1429

Instrument ID: MSVOA_X Calibration Date/Time: 02/26/2025 10:05

Lab File ID: VX045044.D Init. Calib. Date(s): 02/21/2025 02/21/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:58 11:29

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Ethyl Acetate	0.598	0.601		0.5	
Isopropyl Acetate	0.984	1.009		2.54	
Acetone	0.355	0.399		12.39	
Methylene Chloride	2.563	2.401		-6.32	
1,2-Dichloroethane-d4	2.914	2.879	0.01	-1.2	
Toluene-d8	1.660	1.673	0.01	0.78	
4-Bromofluorobenzene	0.561	0.545	0.2	-2.85	
n-amyl Acetate	0.854	0.789		-7.61	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.