

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1429

Client: Tris Pharma, Inc.

Analytical Method: SW624.1

Datafile : VX045045.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0226WBS01	Ethyl Acetate	20	21.1	ug/L	106			68	140	
	Isopropyl Acetate	20	21.3	ug/L	106			70	130	
	Acetone	100	130	ug/L	130			41	148	
	Methylene Chloride	20	19.4	ug/L	97			60	140	
	n-amyl Acetate	20	19.7	ug/L	99			70	130	